

**RULES
OF
TENNESSEE MEDICAL LABORATORY BOARD**

**CHAPTER 1200-06-03
GENERAL RULES GOVERNING MEDICAL LABORATORIES**

TABLE OF CONTENTS

1200-06-03-.01	Definitions	1200-06-03-.12	Reportable Diseases
1200-06-03-.02	Licensing Procedures and Licensure Exemptions	1200-06-03-.13	Personnel Policies
1200-06-03-.03	Change in Location, Director, Owner, Supervisor or Testing in a Medical Laboratory	1200-06-03-.14	Personnel Requirements for a Medical Laboratory
1200-06-03-.04	Inspections - Reports	1200-06-03-.15	Special Regulations for ASTC, Blood Donor Centers, and Plasmapheresis Centers
1200-06-03-.05	Licensure Discipline, Civil Penalties, Assessment of Costs, and Subpoenas	1200-06-03-.16	Alternate Site Testing
1200-06-03-.06	Minimum Standards	1200-06-03-.17	Waived Testing
1200-06-03-.07	Participation In Proficiency Testing	1200-06-03-.18	Collection Stations
1200-06-03-.08	Patient Test Management	1200-06-03-.19	Preparatory Portions of Laboratory Tests
1200-06-03-.09	Quality Control	1200-06-03-.20	Advertising
1200-06-03-.10	Performance Improvement Program	1200-06-03-.21	Renewal of Laboratory License
1200-06-03-.11	Facilities, Safety, Infectious and Hazardous Waste Disposal	1200-06-03-.22	Special Regulations Regarding Laboratories Licensed Under T.C.A. § 68-29-138

1200-06-03-.01 DEFINITIONS. As used in this chapter of Rules the following terms and acronyms shall have the following meanings ascribed to them:

- (1) ASTC - Ambulatory Surgical Treatment Center, specifically any institution, place or building devoted primarily to the maintenance and operation of a facility for the performance of surgical procedures or any facility in which a medical or surgical procedure is utilized to terminate a pregnancy and is licensed by the Tennessee Board for Licensing Health Care Facilities under Rule 1200-8-10.
- (2) Anatomic laboratory - Any medical laboratory performing only the biophysical examination of specimens pertaining to the clinical specialty of pathology, to include histopathology, oral pathology, and cytopathology. The examination of these specimens taken from the human body are performed to obtain information for diagnosis, prophylaxis, or treatment or where any examination, determination or test is made of any sample used as a basis for health advice, or where any sample is collected for the purpose of transfusion or processing of blood or blood fractions, or the training of medical laboratory personnel. Compliance with Rule 1200-6-1-.20 is required to be the director of an anatomic laboratory.
- (3) Authorized person - A physician or intern or resident in an American Medical Association approved training program or a duly licensed optometrist or a duly licensed dentist or a duly licensed chiropractic physician or other health care professional legally permitted to submit to a medical laboratory a written request for tests appropriate to that professional's practice or a law enforcement officer acting in accordance with T.C.A. § 55-10-406.
- (4) Board - The Tennessee Medical Laboratory Board.
- (5) CLIA - The Clinical Laboratory Improvement Amendments as found in 42 CFR 493.
- (6) CLSI – The Clinical Laboratory and Standards Institute.
- (7) CMS - The Centers for Medicare and Medicaid Services of the U.S. Department of Health and Human Services.

(Rule 1200-06-03-.01, continued)

- (8) Center for Disease Control (CDC) - The United States Center for Disease Control, headquartered in Atlanta, GA.
- (9) Clinical Laboratory - see Medical Laboratory.
- (10) Code of Federal Regulations (CFR) - The documents that detail under Title 42, Chapter IV Part 493, the Laboratory Requirements under CLIA, and under Title 21, Parts 600 to 799, the Food and Drug Requirements.
- (11) Condition - A specific requirement from the Board's rules that must be met for continued possession of a license to operate a medical laboratory in Tennessee. Unmet conditions of compliance with the Board's rules can result in harm to patients or facility personnel.
- (12) Condition Level Deficiency - A significant deficiency of non-compliance with a rule. A deficiency that must be corrected before a license is continued, issued or is renewed.
- (13) Days - Calendar days unless otherwise designated.
- (14) Department - The Department of Health, State of Tennessee.
- (15) FDA - The U.S. Food and Drug Administration.
- (16) HCF - The Department of Health, Division of Health Care Facilities.
- (17) HHS - The U.S. Department of Health and Human Services.
- (18) Human Leukocyte Antigens (HLA) - Identification of the human histocompatibility complex antigens on the surface of nucleated cells.
- (19) Independent Lab - A medical laboratory performing patient tests that is independent of both attending and consulting physician's offices and independent of a hospital also known as a reference laboratory.
- (20) Manufacturer's instructions - Information found on the FDA approved product label.
- (21) May - Discretionary.
- (22) Medical Laboratory - Any institution, building, or place in which operations and procedures for the microbiological, serological, chemical, hematological, immunohematological, or biophysical examination of specimens taken from the human body are performed to obtain information for diagnosis, prophylaxis, or treatment or where any examination, determination, or test is made on any sample used as a basis for health advice, or where any sample is collected for the purpose of transfusion or processing of blood fractions, or for the training of medical laboratory personnel.
- (23) Must - Mandatory.
- (24) Office for the Medical Laboratory Board - The Office responsible for laboratory facility licensure.
- (25) Performance Improvement (PI) - An ongoing process encompassing all facets of the laboratory or collection station technical and nontechnical functions at all locations/sites where testing is performed.
- (26) Plan of Correction - The laboratory's or collection station's response to any citation of non-compliance with regulations. The Plan of Correction must be reasonable, contain specific

(Rule 1200-06-03-.01, continued)

actions to correct deficiencies with estimated correction dates and contain the signature of the Laboratory Director or his authorized Designee or collection station director or his authorized designee.

- (27) Point of Care Laboratory Testing - Those tests performed by health care professionals, not licensed by this chapter, upon approval by the board, provided that the testing is performed outside the duly licensed laboratory, and under the auspices of a laboratory required to be licensed by the division of health care facilities in the department of health, pursuant to this chapter.
- (28) Physician Office Laboratory - A laboratory operated by a duly licensed physician for the sole purpose of testing samples collected from their own patients.
- (29) Reference Laboratory - A medical laboratory performing patient tests that is independent of both attending and consulting physician's offices and independent of a hospital also known as an independent laboratory.
- (30) Shall - Mandatory.
- (31) Standard - A specific requirement from the state rules that must be met for continued possession of a license to operate a medical laboratory or collection station in Tennessee. Unmet standards of compliance usually reflect unsatisfactory performance of laboratory or collection station operations according to accepted standards of medical laboratory practice.
- (32) Standard Level Deficiency - A deficiency of non-compliance with a regulation for which a license would be issued or renewed, provided an acceptable plan of correction with an estimated completion date has been submitted and approved.
- (33) Statement of Deficiencies - An official notification of non-compliance with one (1) or more regulatory requirements.
- (34) Unsatisfactory Proficiency Testing Performance - Failure to attain the minimum satisfactory score for an analyte, test, subspecialty, or specialty for a testing event.
- (35) Unsuccessful Proficiency Testing Performance - Failure to attain the minimum satisfactory score for an analyte, test, subspecialty, or specialty for two (2) consecutive or two (2) of three (3) consecutive testing events.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-103, 68-29-105, 63-29-111, 68-29-116, and Public Acts of 2002, Chapter 623. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Amendment filed June 30, 1987; effective August 14, 1987. Repeal and new rule filed May 15, 1990; effective June 29, 1990. Amendment filed November 30, 1990; effective January 14, 1991. Amendment filed February 21, 1991; effective April 7, 1991. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed May 23, 2003; effective August 6, 2003. Amendment filed March 14, 2006; effective May 28, 2006.

1200-06-03-.02 LICENSING PROCEDURES AND LICENSURE EXEMPTIONS.

- (1) Applications
 - (a) An application for a medical laboratory license shall be made under oath, by the owner and director of the medical laboratory or public official responsible for the operation of a city or county medical laboratory or institution that contains a medical laboratory, upon forms provided by the Board.

(Rule 1200-06-03-.02, continued)

- (b) All laboratories performing tests must file a separate application for each laboratory address.
 - (c) Laboratories within a hospital that are located in contiguous buildings on the same campus and under common direction may file a single application or multiple applications for the laboratory sites within the same physical location or street address.
 - (d) Each application shall include, in alphabetical order, a roster of all personnel currently employed in the laboratory, the classification/category in which the employee functions and is licensed, license number, expiration date and social security number. This shall include all medical laboratory directors, consultants, supervisors and testing personnel.
 - (e) The application will be returned to the applicant if it is incomplete or requires correction.
 - (f) After submission of the completed application, the laboratory or collection station may be issued a letter of temporary license which shall be valid for a period of not longer than three (3) months from the date the application is received by the Board. This temporary license will allow the laboratory to perform testing and allow the collection station to collect specimens until an on-site inspection is conducted and a determination is made whether the laboratory or collection station is in compliance. The laboratory shall be fully operational, including but not limited to the reporting of patient results, at the time of the on-site inspection. The collection station shall be fully operational at the time of the on-site inspection. If the laboratory or collection station is found to be in compliance, it shall be issued a permanent license, provided all other requirements for licensure have been met.
- (2) License - Issued
- (a) A license shall be issued authorizing the performance of one (1) or more medical laboratory specialties or subspecialties, the location of testing, the name of the laboratory director and type of ownership.
 - (b) A license shall be issued authorizing the collection of specimens, the location of collection, the name of the physician and type of ownership.
 - (c) A license shall be valid only in the hands of the person or persons to whom it is issued and shall not be the subject of sale, assignment, or transfer, voluntary or involuntary, nor shall a license be valid for any premises other than those for which issued.
 - (d) The license shall be displayed at all times in a prominent place where it may be viewed by the public.
- (3) Fees - The fees established by the Board pursuant to the Tennessee Medical Laboratory Act (T.C.A. §§ 68-29-101, et seq.) for licensure are as follows:
- (a) Initial Laboratory Application Fee - This fee is nonrefundable. \$1000.00
 - (b) Laboratory License Renewal Fee - To be paid annually by all licensees. \$1000.00
 - (c) Late Laboratory License or Collection Station Renewal Penalty Fee - \$ 500.00
To be paid when a licensee fails to timely renew annual license.
 - (d) Annual State Regulatory Fee - To be paid by all licensees. \$ 5.00

(Rule 1200-06-03-.02, continued)

- | | | |
|-----|--|-----------|
| (e) | Replacement License (other than 1200-06-03-.03) when lost or misplaced. | \$ 50.00 |
| (f) | Initial Collection Station Application Fee | \$ 700.00 |
| (g) | Collection Station Renewal Fee | \$ 700.00 |
| (h) | T.C.A. § 68-29-112 Fee – A nonrefundable fee to be paid when there is a change in laboratory ownership, directorship, or location. | \$ 100.00 |
- (4) Fees may be paid in the following manner:
- (a) All fees paid by money order, certified, personal, or corporate check must be submitted to the Board's Administrative Office and made payable to the Tennessee Medical Laboratory Board.
 - (b) Fees may be paid by Division-approved credit cards or other Division-approved electronic methods.
- (5) Pursuant to T.C.A. § 68-29-104(8) the Board hereby exempts the following laboratories (or portions of otherwise licensed laboratories) and the personnel performing tests in those laboratories (or portions of otherwise licensed laboratories) from the requirements imposed pursuant to the "Tennessee Medical Laboratory Act":
- (a) Any laboratory (or portion of an otherwise licensed laboratory) which is operated by the Tennessee Department of Health and in which the only testing on human specimens is that which is required pursuant to an investigation ordered by or in cooperation with the State Epidemiologist or law enforcement agency the results of which are reported only to the State Epidemiologist, CDC or law enforcement officials as required by federal law.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-103, 68-29-104, 68-29-105, and 68-29-113.
Administrative History: Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed June 30, 1987; effective August 14, 1987. Amendment filed February 21, 1991; effective April 7, 1991. Repeal and new rule filed January 7, 1997; effective March 23, 1997. Amendment filed August 31, 2001; effective November 14, 2001. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed July 8, 2004; effective September 21, 2004. Amendment filed December 27, 2006; effective March 12, 2007. Amendment filed April 17, 2007; effective July 1, 2007. Amendment filed April 1, 2009; effective June 15, 2009.

1200-06-03-.03 CHANGE IN LOCATION, DIRECTOR, OWNER, SUPERVISOR OR TESTING IN A MEDICAL LABORATORY.

- (1) It shall be the responsibility of the owner of a laboratory to notify the Department in writing of a change in the location, director, owners or supervisor of the laboratory within fifteen (15) days of the actual change.
- (2) If the matter involves a change of the owner, and/or director and/or the location an application for a new license, including payment of the T.C.A. § 68-29-112 Fee as provided in Rule 1200-06-03-.02 (4), must be filed and a new license obtained before the laboratory may provide services. That new license may be applied for and issued prior to the actual change but will be void should the change not actually take place.
- (3) It shall be the responsibility of the owner to notify the Department in writing in order to add a specialty or subspecialty not presently authorized by the facility's license prior to the

(Rule 1200-06-03-.03, continued)

commencement of testing and reporting patient test results. The Department must conduct an on-site inspection prior to the issuance of authorization for the specialty or subspecialty. A replacement license which includes the new specialty or subspecialty shall be issued by the Department at no cost to the facility.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-105, 68-29-112, and 68-29-114. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed December 27, 2006; effective March 12, 2007.

1200-06-03-.04 INSPECTIONS - REPORTS.

- (1) Biennial Inspections of Licensed Laboratories
 - (a) The Board or its Designee will conduct announced and/or unannounced inspections on at least a biennial basis of any laboratory at any time during its hours of operation to assess compliance with the applicable requirements.
 - (b) It is board policy that regular biennial inspections be unannounced.
 - (c) Written reports of inspection results will be submitted to the medical laboratory director within ten (10) working days of the on-site visit.
 - (d) The medical laboratory director must respond to any deficiencies cited with a written plan of correction that includes estimated correction dates, within ten (10) working days of the receipt of the statement of deficiencies.
- (2) Inspection for Initial License - The laboratory shall be issued an initial license only after an on-site inspection is conducted and the laboratory has no deficiencies, or has submitted an acceptable plan of correction for all standard level deficiencies. All condition level deficiencies must be corrected before the laboratory shall be issued a license.

Authority: T.C.A. §§ 4-5-204, 4-5-204, 68-29-105, 68-29-106, and 68-29-113. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Repeal and new rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002.

RULE 1200-06-03-.05 LICENSURE DISCIPLINE, CIVIL PENALTIES, ASSESSMENT OF COSTS, SUBPOENAS, AND SCREENING PANELS.

- (1) Upon a finding by the Board that a licensee has violated any provision of the Tennessee Medical Laboratory Act (T.C.A. §§ 68-29-101, et seq.) or the rules promulgated pursuant thereto, the Board may impose any of the following actions separately or in any combination which is deemed appropriate to the offense:
 - (a) Private Censure - This is a written action issued by the Board to the licensee for minor infractions of these rules. It is informal and advisory in nature and does not constitute a formal disciplinary action.
 - (b) Public Censure or Reprimand - This is a formal disciplinary action consisting of a written admonishment issued by the Board to a licensee for one (1) time and less severe infractions.
 - (c) Probation - This is a formal disciplinary action which places a licensee on close scrutiny for a fixed period of time. This action may be combined with conditions which

(Rule 1200-06-03-.05, continued)

must be met before probation will be lifted and/or which restrict the licensee's activities during the probationary period.

- (d) Licensure Suspension - This is a formal disciplinary action which suspends a licensee's right to operate a medical laboratory for a fixed period of time. It contemplates the reentry of the licensee into operation under the license previously issued.
 - (e) Licensure Revocation - This is the most severe form of disciplinary action which removes a licensee from the operation of a medical laboratory and terminates the license previously issued. If revoked, it relegates the violator to the status he or she possessed prior to application for licensure. However, the Board may in its discretion allow the reinstatement of a revoked license upon conditions and after a period of time it deems appropriate. No petition for reinstatement and no new application for licensure from a medical laboratory whose license was revoked shall be considered prior to the expiration of at least one (1) year unless otherwise stated in the Board's revocation Order.
 - (f) Civil Penalty - A monetary disciplinary action assessed by the Board pursuant to paragraph (5) of this rule.
 - (g) Once ordered, probation, suspension, revocation, or any other condition of any type of disciplinary action may not be lifted unless and until the licensee petitions, pursuant to paragraph (2) of this rule, and appears before the Board after the period of initial probation, suspension, revocation, or other conditioning has run and all conditions placed on the probation, suspension, revocation, have been met.
- (2) Order of Compliance - This procedure is a necessary adjunct to previously issued disciplinary orders and is available only when a petitioner has completely complied with the provisions of a previously issued disciplinary order, and wishes or is required to obtain an order reflecting that compliance.
- (a) The Board will entertain petitions for an Order of Compliance as a supplement to a previously issued order upon strict compliance with the procedures set forth in subparagraph (b) in only the following three (3) circumstances:
 - 1. When the petitioner can prove compliance with all the terms of the previously issued order and is seeking to have an order issued reflecting that compliance; or
 - 2. When the petitioner can prove compliance with all the terms of the previously issued order and is seeking to have an order issued lifting a previously ordered suspension or probation; or
 - 3. When the petitioner can prove compliance with all the terms of the previously issued order and is seeking to have an order issued reinstating a license previously revoked.
 - (b) Procedures
 - 1. The petitioner shall submit a Petition for Order of Compliance, as contained in subparagraph (c), to the Board's Administrative Office that shall contain all of the following:
 - (i) A copy of the previously issued order; and

(Rule 1200-06-03-.05, continued)

- (ii) A statement of which provision of subparagraph (a) the petitioner is relying upon as a basis for the requested order; and
 - (iii) A copy of all documents that prove compliance with all the terms or conditions of the previously issued order. If proof of compliance requires testimony of an individual(s), including that of the petitioner, the petitioner must submit signed statements from every individual the petitioner intends to rely upon attesting, under oath, to the compliance. The Board's consultant and administrative staff, in their discretion, may require such signed statements to be notarized. No documentation or testimony other than that submitted will be considered in making an initial determination on, or a final order in response to, the petition.
2. The Board authorizes its consultant and administrative staff to make an initial determination on the petition and take one of the following actions:
 - (i) Certify compliance and have the matter scheduled for presentation to the Board as an uncontested matter; or
 - (ii) Deny the petition, after consultation with legal staff, if compliance with all of the provisions of the previous order is not proven and notify the petitioner of what provisions remain to be fulfilled and/or what proof of compliance was either not sufficient or not submitted.
 3. If the petition is presented to the Board the petitioner may not submit any additional documentation or testimony other than that contained in the petition as originally submitted.
 4. If the Board finds that the petitioner has complied with all the terms of the previous order an Order of Compliance shall be issued.
 5. If the petition is denied either initially by staff or after presentation to the Board and the petitioner believes compliance with the order has been sufficiently proven the petitioner may, as authorized by law, file a petition for a declaratory order pursuant to the provisions of T.C.A. § 4-5-223 and rule 1200-10-01-.11.
- (c) Form Petition

Petition for Order of Compliance
Tennessee Medical Laboratory Board

Petitioner's Name: _____

Petitioner's Mailing Address: _____

Petitioner's E-Mail Address: _____

Telephone Number: _____

Attorney for Petitioner: _____

Attorney's Mailing Address: _____

Attorney's E-Mail Address: _____

Telephone Number: _____

(Rule 1200-06-03-.05, continued)

The petitioner respectfully represents, as substantiated by the attached documentation, that all provisions of the attached disciplinary order have been complied with and I am respectfully requesting: (circle one)

1. An order issued reflecting that compliance; or
2. An order issued reflecting that compliance and lifting a previously ordered suspension or probation; or
3. An order issued reflecting that compliance and reinstating a license previously revoked.

Note - You must enclose all documents necessary to prove your request including a copy of the original order. If any of the proof you are relying upon to show compliance is the testimony of any individual, including yourself, you must enclose signed statements from every individual you intend to rely upon attesting, under oath, to the compliance. The Board's consultant and administrative staff, in their discretion, may require such signed statements to be notarized. No documentation or testimony other than that submitted will be considered in making an initial determination on, or a final order in response to, this petition.

Respectfully submitted this the _____ day of _____, 20

Petitioner's Signature

- (3) Order Modifications - This procedure is not intended to allow anyone under a previously issued disciplinary order, to modify any findings of fact, conclusions of law, or the reasons for the decision contained in the order. It is also not intended to allow a petition for a lesser disciplinary action. All such provisions of Board orders were subject to reconsideration and appeal under the provisions of the Uniform Administrative Procedures Act (T.C.A. §§ 4-5-301, et seq.). This procedure is not available as a substitute for reconsideration and/or appeal and is only available after all reconsideration and appeal rights have been either exhausted or not timely pursued. It is also not available for those who have accepted and been issued a reprimand.
- (a) The Board will entertain petitions for modification of the disciplinary portion of previously issued orders upon strict compliance with the procedures set forth in subparagraph (b) only when the petitioner can prove that compliance with any one or more of the conditions or terms of the discipline previously ordered is impossible. For purposes of this rule the term "impossible" does not mean that compliance is inconvenient or impractical for personal, financial, scheduling or other reasons.
- (b) Procedures
1. The petitioner shall submit a written and signed Petition for Order Modification on the form contained in subparagraph (c) to the Board's Administrative Office that shall contain all of the following:
 - (i) A copy of the previously issued order; and
 - (ii) A statement of why the petitioner believes it is impossible to comply with the order as issued; and

(Rule 1200-06-03-.05, continued)

- (iii) A copy of all documents that proves that compliance is impossible. If proof of impossibility of compliance requires testimony of an individual(s), including that of the petitioner, the petitioner must submit signed and notarized statements from every individual the petitioner intends to rely upon attesting, under oath, to the reasons why compliance is impossible. No documentation or testimony other than that submitted will be considered in making an initial determination on, or a final order in response to, the petition.
 - 2. The Board authorizes its consultant and administrative staff to make an initial determination on the petition and take one of the following actions:
 - (i) Certify impossibility of compliance and forward the petition to the Office of General Counsel for presentation to the Board as an uncontested matter; or
 - (ii) Deny the petition, after consultation with legal staff, if impossibility of compliance with the provisions of the previous order is not proven and notify the petitioner of what proof of impossibility of compliance was either not sufficient or not submitted.
 - 3. If the petition is presented to the Board the petitioner may not submit any additional documentation or testimony other than that contained in the petition as originally submitted.
 - 4. If the petition is granted a new order shall be issued reflecting the modifications authorized by the Board that it deemed appropriate and necessary in relation to the violations found in the previous order.
 - 5. If the petition is denied either initially by staff or after presentation to the Board and the petitioner believes impossibility of compliance with the order has been sufficiently proven the petitioner may, as authorized by law, file a petition for a declaratory order pursuant to the provisions of T.C.A. § 4-5-223 and rule 1200-10-01-.11.
- (c) Form Petition

Petition for Order Modification
Tennessee Medical Laboratory Board

Petitioner's Name: _____
 Petitioner's Mailing Address: _____

 Petitioner's E-Mail Address: _____
 Telephone Number: _____

 Attorney for Petitioner: _____
 Attorney's Mailing Address: _____

 Attorney's E-Mail Address: _____
 Telephone Number: _____

(Rule 1200-06-03-.05, continued)

The petitioner respectfully represents that for the following reasons, as substantiated by the attached documentation, the identified provisions of the attached disciplinary order are impossible for me to comply with:

Note – You must enclose all documents necessary to prove your request including a copy of the original order. If any of the proof you are relying upon to show impossibility is the testimony of any individual, including yourself, you must enclose signed and notarized statements from every individual you intend to rely upon attesting, under oath, to the reasons why compliance is impossible. No documentation or testimony other than that submitted will be considered in making an initial determination on, or a final order in response to, this petition.

respectfully submitted this the ____ day of _____, 20____.

Petitioner’s Signature

(4) The provisions of the Uniform Administrative Procedures Act, T.C.A. §§ 4-5-301, et seq., shall govern the hearing and judicial review of all disciplinary proceedings.

(5) Civil Penalties

(a) Purpose - The purpose of this is to set out a schedule designating the minimum and maximum Civil Penalties which may be assessed pursuant to T.C.A. §63-1-134.

(b) Schedule of Civil Penalties

1. A Type A Civil Penalty may be imposed whenever the Board finds the medical laboratory or collection station which is required to be licensed by the Board guilty of a willful and knowing violation of the Tennessee Medical Laboratory Act, or regulations promulgated pursuant thereto, to such an extent that there is, or is likely to be an imminent, substantial threat to the health, safety and welfare of an individual patient or the public. For purposes of this section, willfully and knowingly operating a medical laboratory or collection station without a license from the Board is one of the violations of the Tennessee Medical Laboratory Act for which a Type A Civil Penalty is assessable.

2. A Type B Civil Penalty may be imposed whenever the Board finds the medical laboratory or collection station which is required to be licensed by the Board is guilty of a violation of the Tennessee Medical Laboratory Act, or regulations promulgated pursuant thereto, in such a manner as to impact directly on the care of patients or the public.

3. A Type C Civil Penalty may be imposed whenever the Board finds the medical laboratory or collection station which is required to be licensed by the Board is guilty of a violation of the Tennessee Medical Laboratory Act, or regulations promulgated pursuant thereto, which is neither directly detrimental to the patients or public, nor directly impacts their care, but has only indirect relationship to patient care or the public.

(c) Amount of Civil Penalties

(Rule 1200-06-03-.05, continued)

1. Type A Civil Penalties shall be assessed in the amount of not less than \$500 nor more than \$1,000.
 2. Type B Civil Penalties may be assessed in the amount of not less than \$100 and not more than \$500.
 3. Type C Civil Penalties may be assessed in the amount of not less than \$50 and not more than \$100.
- (d) Procedures for Assessing Civil Penalties
1. The Division of Health Related Boards may initiate a Civil Penalty assessment by filing a Memorandum of Assessment of Civil Penalty. The Division shall state in the memorandum the facts and law upon which it relies in alleging a violation, the proposed amount of the civil penalty and the basis for such penalty. The Division may incorporate the Memorandum of Assessment of Civil Penalty with a Notice of Charges which may be issued attendant thereto.
 2. Civil Penalties may also be initiated and assessed by the Board during consideration of any Notice of Charges. In addition, the Board may, upon good cause shown, assess a type and amount of civil penalty which was not recommended by the Division.
 3. In assessing the Civil Penalties pursuant to these rules the Board may consider the following factors:
 - (i) Whether the amount imposed will be a substantial economic deterrent to the violator;
 - (ii) The circumstances leading to the violation;
 - (iii) The severity of the violation and the risk of harm to the public;
 - (iv) The economic benefits gained by the violator as a result of non-compliance; and,
 - (v) The interest of the public.
 4. All proceedings for the assessment of Civil Penalties shall be governed by the contested case provisions of Title 4, Chapter 5, Tennessee Code Annotated.
- (6) Assessment of costs in disciplinary proceedings shall be as set forth in T.C.A. §§ 63-1-144 and 68-29-136.
- (7) Subpoenas
- (a) Purpose - Although this rule applies to persons and entities other than medical laboratories, it is the Board's intent as to medical laboratories that they be free to operate without fear that such operation or its documentation will be unduly subjected to scrutiny outside the profession. Consequently, balancing that intent against the interest of the public and patients to be protected against substandard care and activities requires that persons seeking to subpoena such information and/or materials must comply with the substance and procedures of these rules.

It is the intent of the Board that the subpoena power outlined herein shall be strictly proscribed. Such power shall not be used by the Division or Board investigators to seek other incriminating evidence against medical laboratory personnel when the

(Rule 1200-06-03-.05, continued)

Division or Board does not have a complaint or basis to pursue such an investigation. Thus, unless the Division or its investigators have previously considered, discovered, or otherwise received a complaint from either the public or a governmental entity, then no subpoena as contemplated herein shall issue.

(b) Definitions - As used in this chapter of rules the following words shall have the meanings ascribed to them:

1. Probable Cause

(i) For investigative subpoenas shall mean that probable cause, as defined by case law at the time of request for subpoena issuance is made, exists that a violation of the Medical Laboratory Practice Act or rules promulgated pursuant thereto has occurred or is occurring and that it is more probable than not that the person(s), or items to be subpoenaed possess or contain evidence which is more probable than not relevant to the conduct constituting the violation.

(ii) The utilization of the probable cause evidentiary burden in proceedings pursuant to this rule shall not in any way, nor should it be construed in any way to establish a more restrictive burden of proof than the existing preponderance of the evidence in any civil disciplinary action which may involve the person(s) or items that are the subject of the subpoena.

2. Presiding Officer - For investigative subpoenas shall mean any elected officer of the Board, or any duly appointed or elected chairperson of any panel of the Board.

(c) Procedures

1. Investigative Subpoenas

(i) Investigative Subpoenas are available only for issuance to the authorized representatives of the Tennessee Department of Health, its investigators and its legal staff.

(ii) An applicant for such a subpoena must either orally or in writing notify the Board's Unit Director of the intention to seek issuance of a subpoena. That notification must include the following:

(I) The time frame in which issuance is required so the matter can be timely scheduled; and

(II) A particular description of the material or documents sought, which must relate directly to an ongoing investigation or contested case, and shall, in the instance of documentary materials, be limited to the records of the patient or patients whose complaint, complaints, or records are being considered by the Division or Board.

(III) In no event shall such subpoena be broadly drafted to provide investigative access to medical laboratory records of other patients who are not referenced in a complaint received from an individual or governmental entity, or who have not otherwise sought relief, review, or Board consideration of any medical laboratory personnel's conduct, act, or omission; and

(Rule 1200-06-03-.05, continued)

- (IV) Whether the proceedings for the issuance is to be conducted by physical appearance or electronic means; and
 - (V) The name and address of the person for whom the subpoena is being sought, or who has possession of the items being subpoenaed.
- (iii) The Board's Unit Director shall cause the following to be done:
- (I) In as timely a manner as possible arrange for either an elected officer of the Board or any duly appointed or elected chairperson of any panel of the board to preside and determine if issuing the subpoena should be recommended to the full board; and
 - (II) Establish a date, time and place for the proceedings to be conducted and notify the Presiding Officer, the applicant and the court reporter; and
 - (III) Maintain a complete record of the proceedings including an audio tape in such a manner as to:
 - I. Preserve a verbatim record of the proceeding; and
 - II. Prevent the person presiding over the proceedings and/or signing the subpoena from being allowed to participate in any manner in any resulting disciplinary action of any kind, formal or informal, which involves either the person or the documents or records for which the subpoena was issued.
- (iv) The Proceedings
- (I) The applicant shall do the following:
 - I. Provide for the attendance of all persons whose testimony is to be relied upon to establish probable cause; and
 - II. Produce and make part of the record copies of all documents to be utilized to establish probable cause; and
 - III. Obtain, complete and provide to the Presiding Officer a subpoena which specifies the following:
 - A. The name and address of the person for whom the subpoena is being sought or who has possession of the items being subpoenaed; and
 - B. The location of the materials, documents or reports for which production pursuant to the subpoena is sought if that location is known; and
 - C. A brief, general description of any materials, documents or items to be produced pursuant to the subpoena; and
 - D. The date, time and place for compliance with the subpoena.

(Rule 1200-06-03-.05, continued)

- IV. Provide the Presiding Officer testimony and/or documentary evidence which, in good faith, the applicant believes is sufficient to establish that probable cause exists for issuance of the subpoena as well as sufficient proof that all other reasonably available alternative means of securing the materials, documents or items have been unsuccessful.

(II) The Presiding Officer shall do the following:

- I. Be selected only after assuring the Board's Unit Director that he or she has no prior knowledge of or any direct or indirect interest in or relationship with the person(s) being subpoenaed and/or the licensee who is the subject of the investigation; and
- II. Commence the proceedings and swear all necessary witnesses; and
- III. Hear and maintain the confidentiality, if any, of the evidence presented at the proceedings and present to the full board only that evidence necessary for an informed decision; and
- IV. Control the manner and extent of inquiry during the proceedings and be allowed to question any witness who testifies; and
- V. Determine, based solely on the evidence presented in the proceedings, whether probable cause exists and if so, make such recommendation to the full Board; and
- VI. Not participate in any way in any other proceeding whether formal or informal, which involves the matters, items or person(s) which are the subject of the subpoena. This does not preclude the Presiding Officer from presiding at further proceedings for issuance of subpoenas in the matter.

(III) The Board shall do the following:

- I. By a vote of two thirds (2/3) of the Board members issue the subpoena for the person(s) or items specifically found to be relevant to the inquiry, or quash or modify an existing subpoena by a majority vote; and
- II. Sign the subpoena as ordered to be issued, quashed or modified.

2. Post-Notice of Charges Subpoenas - If the subpoena is sought for a contested case being heard with an Administrative Law Judge from the Secretary of State's office presiding, this definition shall not apply and all such post-notice of charges subpoenas should be obtained from the office of the Administrative Procedures Division of the Office of the Secretary of State pursuant to the Uniform Administrative Procedures Act and rules promulgated pursuant thereto.

(d) Subpoena Forms

1. All subpoenas shall be issued on forms approved by the Board.

(Rule 1200-06-03-.05, continued)

2. The subpoena forms may be obtained by contacting the Board's Administrative Office.
- (e) Subpoena Service - Any method of service of subpoenas authorized by the Tennessee Rules of Civil Procedure or the rules of procedure for contested cases of the Tennessee Department of State, Administrative Procedures Division may be utilized to serve subpoenas pursuant to this rule.
- (8) Screening Panels - The Board adopts, as if fully set out herein, rule 1200-10-01-.13, of the Division of Health Related Boards and as it may from time to time be amended, as its rule governing the screening panel process.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 4-5-217, 4-5-223, 4-5-311, 63-1-122, 63-1-134, 63-1-138, 63-1-144, 68-29-104, 68-29-105, 68-29-109, 68-29-111, 68-29-122, 68-29-126, 68-29-128, 68-29-129, 68-29-130, 68-29-131, 68-29-132, 68-29-136, and 68-29-405. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Amendment filed September 30, 1987; effective November 14, 1987. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Amendment filed December 1, 2000; effective February 14, 2001. Amendment filed March 22, 2001; effective June 5, 2001. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed July 8, 2004; effective September 21, 2004. Amendments filed May 16, 2005; effective July 30, 2005. Amendment filed March 16, 2007; effective May 30, 2007.

1200-06-03-.06 MINIMUM STANDARDS. Each laboratory shall comply with the standards as set forth in 42 C. F. R. Part 493 of the current CLIA Regulations. These standards are deemed to be the minimum standard acceptable to the Board. Failure to meet the CLIA standards may result in disciplinary action as defined in Rule 1200-06-03-.05.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, 68-29-105, and 68-29-126. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.07 PARTICIPATION IN PROFICIENCY TESTING.

- (1) Condition: Enrollment and Testing of Samples - Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in 42 CFR Part 493, Subpart I of the current CLIA Regulations and is approved by HHS. The laboratory must enroll in an approved program for each of the specialties and subspecialties for which it seeks licensure. The laboratory must test the samples in the same manner as patient's specimens.
 - (a) Standard Enrollment.
 1. The laboratory must ensure that the Board is notified yearly of the approved program or programs in which it chooses to participate to meet proficiency testing requirements of this rule.
 2. The laboratory must designate the program(s) to be used for each specialty, subspecialty, and analyte or task to determine compliance with the rule if the Laboratory participates in more than one (1) proficiency testing program approved by CMS.
 3. For those tests performed by the laboratory for which proficiency testing is not required by 42 CFR Part 493.901, Subpart H of the CLIA regulations, a laboratory must establish and maintain the accuracy and reliability of its testing procedures, in accordance with Rule 1200-06-03-.10 (5).

(Rule 1200-06-03-.07, continued)

4. For each specialty, subspecialty and analyte or test, the laboratory must participate in one (1) approved proficiency testing program or programs, for one (1) year before designating a different program and must notify the Board before any change in designation.
 5. The laboratory must authorize the proficiency testing program to release to the Board all data that may be required to determine the laboratory's compliance with this rule.
- (b) Standard: Testing of proficiency testing samples. The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens.
1. The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods.
 2. The individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient workload using the laboratory's routine methods.
 3. The laboratory must test samples the same number of times that it routinely test patient samples.
 4. Laboratories that perform tests on proficiency testing samples must not engage in any inter-laboratory communications pertaining to the results of proficiency testing sample(s) until after the date by which the laboratory must report proficiency testing results to the program for the testing event in which the samples were sent.
 5. Laboratories with multiple testing sites or separate locations must not participate in any communications or discussions across sites/locations concerning proficiency testing sample results until after the date by which the laboratory must report proficiency testing results to the program.
 6. The laboratory must not send proficiency testing samples or portions of samples to another laboratory for any analysis which they are certified to perform in their own laboratory.
 7. Any laboratory that the Board determines intentionally referred its proficiency testing samples to another laboratory for analysis and submits the other laboratory's results as their own shall have its license disciplined, pursuant to Rule 1200-06-03-.05.
 8. Any laboratory that receives proficiency testing samples from another laboratory for testing must notify the Board of the receipt of those samples.
 9. The laboratory must document the handling, preparation, processing, examination, and each step in the testing and reporting of results for all proficiency testing samples.
 10. The laboratory must maintain copy of all records, including a copy of the proficiency testing program report forms used by the laboratory to record proficiency testing results including the attestation statement provided by the proficiency testing program, signed by the analyst and the laboratory director,

(Rule 1200-06-03-.07, continued)

documenting that proficiency testing samples were tested in the same manner as patient specimens, for a minimum of two (2) years from the date of the proficiency testing event.

11. Unless otherwise specified in the Rules, proficiency testing is required for only the test system, assay, or examination used as the primary method for patient testing during the proficiency testing event.

(2) Condition: Successful Participation:

- (a) Each laboratory performing tests must successfully participate in a proficiency testing program approved by CMS, if applicable, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA.
- (b) If the laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test sanctions will be taken as defined in Rule 1200-06-03-.05.

(3) Condition: Reinstatement of Laboratories after Failure to Participate Successfully.

- (a) If a laboratory's license is suspended or placed on probation by the Board or limited pursuant to CLIA regulations because it fails to participate successfully in proficiency testing for one (1) or more specialties, subspecialties, analyte, or voluntarily surrenders its license for the failed specialty, subspecialty, or analyte, the laboratory must then demonstrate sustained satisfactory performance on two (2) consecutive proficiency testing events, before the Board will consider it for reinstatement in that specialty, subspecialty, analyte or test.
- (b) The period of suspension of licensure for the failed specialty, subspecialty, or analyte or test is for a period of not less than six (6) months from the date of suspension.
- (c) If a laboratory's license is suspended, the laboratory must take corrective action and petition the Board for the suspension to be lifted and apply for the license to be reinstated and pay any necessary fees as set out in 1200-06-03-.02.

(4) Condition: Specialties and Subspecialties.

- (a) Failure to attain an overall testing event score of at least eighty per cent (80%) is unsatisfactory performance for Bacteriology, Mycobacteriology, Mycology, Parasitology, Virology, Syphilis Serology, General Immunology, Routine Chemistry, Endocrinology, Toxicology, Hematology, Antibody Detection, and Anti-Body Identification.
- (b) Failure to attain a score of at least eighty per cent (80%) of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event of General Immunology, Routine Chemistry, Endocrinology, Toxicology, and Hematology.
- (c) Failure to attain an overall testing event score of at least one hundred per cent (100%) is unsatisfactory performance for ABO grouping, Rh typing, and Compatibility testing.
- (d) Failure to participate in a testing event is unsatisfactory performance and results in a score of zero (0) for the testing event.
- (e) Consideration may be given to those laboratories failing to participate in a testing event only if the following requirements are met:

(Rule 1200-06-03-.07, continued)

1. Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; and
 2. The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and
 3. The laboratory had successfully participated in the previous two (2) proficiency testing events.
- (f) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of zero (0) for the testing event.
- (g) For any unsatisfactory analyte or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure.
- (h) For any unsatisfactory analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two (2) years from the date of participation in the proficiency testing event.
- (i) Failure to achieve satisfactory performance for the same analyte or test in two (2) consecutive testing events or two (2) out of three (3) consecutive testing events is unsuccessful performance.
- (j) Failure to achieve an overall testing event score of satisfactory performance for two (2) consecutive testing events or two (2) out of three (3) consecutive testing events is unsuccessful performance.
- (k) All unsuccessful testing events shall be reported to the Department within thirty (30) days of receipt of the hard copy print-out that documents results for unsuccessful proficiency testing.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Amendment filed September 30, 1987; effective November 14, 1987. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed August 25, 2003; effective November 8, 2003. Amendment filed October 11, 2004; effective December 25, 2004.

1200-06-03-.08 PATIENT TEST MANAGEMENT.

- (1) Condition: Each laboratory must employ and maintain a system that provides for proper patient preparation; proper specimen collection, identification, preservation, transportation, and processing; and accurate result reporting. This system must assure optimum patient specimen integrity and positive identification throughout the testing process and must meet the standards of this rule as they apply to the testing performed.
- (a) Standard: Procedures for specimen submission and handling:
1. The laboratory must have available and must follow written policies and procedures for each of the following, if applicable: methods used for the

(Rule 1200-06-03-.08, continued)

- preparation of patients; specimen collection; specimen labeling; specimen preservation; and conditions for specimen transportation and specimen processing.
2. Such policies and procedures must assure positive identification and optimum integrity of patient specimens from the time the specimens are collected until testing has been completed and the results reported.
 3. If the laboratory accepts referral specimens, written instructions must be available to clients and must include, as appropriate, the information specified in paragraph (a) 1. of this section.
 4. Oral explanation of instructions to patients for specimen collections, including patient preparation, may be used as a supplement to written instructions where applicable.
- (b) Standard: Test Requisition.
1. The laboratory must perform tests only at the written or electronic request of an authorized person. The laboratory must maintain the written authorization or documentation of efforts made to obtain a written authorization.
 2. Outpatient laboratory testing in Tennessee hospitals may be ordered by the following:
 - (i) Any licensed Tennessee practitioner who is authorized to do so by T.C.A. § 68-29-121;
 - (ii) Any out of state practitioner who has a Tennessee telemedicine license issued pursuant to rule 0880-02-.16; or
 - (iii) Any duly licensed out of state health care professional as listed in T.C.A. § 68-29-121 who is authorized by his or her state board to order outpatient laboratory testing in hospitals for individuals with whom that practitioner has an existing face-to-face patient relationship as outlined in rule 0880-02-.14(7)(a)1., 2., and 3.
 3. Records of test requisitions or test authorizations must be retained for a minimum of two (2) years. The patient's chart or medical record, if used as the test requisition, must be retained for a minimum of two (2) years and must be available to the laboratory at the time of testing and available to the Board upon request.
 4. The laboratory must assure that the requisition or test authorization includes all of the following:
 - (i) The patient's name or other unique identifier.
 - (ii) The name and address or other suitable identifiers of the authorized person requesting the test and, if appropriate, the individual responsible for utilizing the test results or the name and address of the laboratory submitting the specimen, including, as applicable, a contact person to enable the reporting of imminent life threatening laboratory results or panic values;
 - (iii) The test(s) to be performed;

(Rule 1200-06-03-.08, continued)

- (iv) The date of specimen collection;
- (v) For PAP smears, the patient's last menstrual period, age or date of birth, and indication of whether the patient had a previous abnormal report, treatment or biopsy, and
- (vi) Any additional information relevant and necessary to a specific test to assure accurate and timely testing and reporting of results.

(c) Standard: Test Records:

1. The laboratory must maintain a record system to ensure reliable identification of patient specimens as they are processed and tested to assure that accurate test results are reported. These records must identify the personnel performing the testing procedure.
2. Records of patient testing, including, if applicable, instrument printouts or electronic data storage, must be retained for at least two (2) years. Immunohematology and transfusion records must be retained for no less than five (5) years.
3. The record system must provide documentation of information specified in Rule 1200-06-03-.08(1)(b)3. (i) through (vi) and include all of the following:
 - (i) The patient identification number, accession number, or other unique identification of the specimen;
 - (ii) The date and time of specimen receipt into the laboratory;
 - (iii) The condition and disposition of specimens that do not meet the laboratory's criteria for specimen acceptability and
 - (iv) The records and dates of all specimen testing, including the identity of the personnel who performed the test(s), which are necessary to assure proper identification and accurate reporting of patient test results.

(d) Standard: Test report.

1. The results of a laboratory test shall be reported promptly to the authorized person, the individual responsible for using the test results, the laboratory initially requesting the test, or other person required by statute or rule to receive test results.
2. The original report or an exact duplicate of each test report, including final and preliminary reports, must be retained by the testing laboratory for a period of at least two (2) years after the date of reporting. This information may be maintained as part of the patient's chart or medical record which must be readily available to the laboratory and to the Board upon request.
 - (i) Immunohematology reports must be retained by the laboratory for a period of no less than five (5) years.
 - (ii) Pathology test reports must be retained for a period of at least ten (10) years after the date of reporting.

(Rule 1200-06-03-.08, continued)

3. The laboratory must have adequate systems in place to report results in a timely, accurate, reliable and confidential manner, and ensure patient confidentiality throughout those parts of the total process that are under the laboratory's control.
4. Laboratories must comply with Rule 1200-14-01-.41, Reports of Sexually Transmitted Diseases, which states, in part, that the following diseases are declared to be sexually transmitted diseases and, upon their diagnosis or treatment, are subject to reporting requirements as designated in T.C.A. § 68-10-101.
 - (i) Acquired Immune Deficiency Syndrome (AIDS) I
 - (ii) Gonorrhea I
 - (iii) Syphilis (by stage) I
 - (iv) Chlamydia trachomatis
 - (v) Nongonococcal Urethritis/NGU (Number of Cases)
 - (vi) Human Immunodeficiency Virus - all types

Confidential (opaque envelope) report required. All information and reports concerning persons infected with sexually transmitted diseases shall be confidential and shall be inaccessible to the public.

5. The test report must indicate the name and address of the laboratory location at which the test was performed, the test performed, the name of the medical laboratory director, the test result and, if applicable, the units of measurement.
6. The laboratory must indicate on the test report any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.
7. Pertinent "reference, or "normal" ranges, as determined by the laboratory performing the tests must be available to the authorized person who ordered the tests or the individual responsible for utilizing the test results.
8. The results or transcripts of laboratory tests or examinations must be released only to authorized persons or the individual responsible for utilizing the test results.
9. The laboratory must develop and follow written procedures for reporting imminent life-threatening laboratory results or "critical values." In addition, the laboratory must immediately alert the individual, entity requesting the test or the individual responsible for utilizing the test results.
10. The laboratory must, upon request, make available to clients a list of test methods employed by the laboratory and, in accordance with Rule 1200-06-03-.09 (4), as applicable, the performance specifications of each method used to test patient specimens. In addition, information that may affect the interpretation of test results, such as test interference must be provided upon request.
11. Pertinent updates on testing information must be provided to clients whenever changes occur that affect the test results or interpretation of test results.

(Rule 1200-06-03-.08, continued)

12. The original report or exact duplicates of test reports must be maintained by the laboratory in a manner that permits ready identification and timely accessibility.
- (e) Standard: Referral of specimens - A laboratory must refer specimens for testing only to an in state laboratory possessing a current Tennessee license and current CLIA certification, or an out of state laboratory having a CLIA certification authorizing the performance of testing in the specialty or subspecialty of service for the level of complexity in which the referred test is categorized. The laboratory shall maintain a copy of the current CLIA certification for the referral or reference laboratory.
1. The referring laboratory must not revise results or information directly related to the interpretation of results provided by the testing laboratory.
 2. The referring laboratory may permit each testing laboratory to send the test result directly to the authorized person who initially requested the test. The referring laboratory must retain or be able to produce an exact duplicate of each testing laboratory's report.
 3. The authorized person who orders a test or procedure must be notified by the referring laboratory of the name and address of each laboratory location at which a test was performed.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-103, 68-29-105, 68-29-107, 68-29-121, and 68-29-124.
Administrative History: Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed October 11, 2004; effective December 25, 2004. Amendment filed August 8, 2005; effective October 22, 2005.

1200-06-03-.09 QUALITY CONTROL.

- (1) Condition: General Quality Control - The laboratory must establish and follow written quality control procedures, including but not limited to electronic quality control and internal quality control, for monitoring and evaluating the quality of the analytical testing process of each method. General quality control shall assure the accuracy and reliability of patient test results and reports. The medical laboratory director shall review all general quality control or may designate, in writing, the review to personnel meeting the qualification of those respective positions. The medical laboratory director shall retain the ultimate responsibility for the general quality control of the operation of the laboratory. In addition, the laboratory must meet the applicable standards in paragraphs 1200-06-03-.09(2) through 1200-06-03-.09(11) of this rule.
- (2) Standard: Laboratory Testing: For each test performed, the laboratory will be in compliance if it
 - (a) Meets all applicable quality control requirements specified in this rule; or
 - (b) Follows manufacturer's instructions when using their products (instruments, kits, or test systems). In addition, the laboratory must comply with requirements within any paragraph of this rule that are unique to the laboratory facility and cannot be met by manufacturer's instructions.
- (3) Standard: Test Methods, Equipment, Instrumentation, Reagents, Materials, and Supplies. The laboratory must utilize test methods, equipment instrumentation, reagents, materials, and supplies that provide accurate and reliable test results and test reports.

(Rule 1200-06-03-.09, continued)

- (a) Test methodologies and equipment must be selected and testing performed in a manner that provides test results within the laboratory's stated performance specifications for each test method as determined under Rule 1200-06-03-.09(4).
 1. The laboratory must have appropriate and sufficient equipment, instruments, reagents, materials, and supplies for the type and volume of testing performed and for the maintenance of quality during all phases of testing.
 2. The accuracy of analytical balance weights must be verified annually against appropriate standard sources.
 3. The accuracy of thermometers must be verified annually against appropriate standard sources or following the manufacturer's instructions.
 4. The accuracy of pipettes must be verified annually against appropriate standard sources or following the manufacturer's instructions.
 5. The accuracy of centrifuge(s) must be verified annually against appropriate standard sources or following manufacturer's instructions.
 6. The accuracy of timers must be verified annually against, appropriate standard sources or following manufacturer's instructions.
- (b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, for accurate and reliable test system operation, and for test result reporting.
 1. These conditions include, if applicable, the following:
 - (i) Water quality;
 - (ii) Temperature;
 - (iii) Humidity; and
 - (iv) Protection of equipment and instrumentation from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.
 2. Remedial actions taken to correct conditions that fail to meet the criteria specified in this subparagraph must be documented.
- (c) Reagents, solutions, culture media, control materials, calibration materials and other supplies, as appropriate, must be labeled to indicate the following:
 1. Identity and, when significant, titer, strength or concentration;
 2. Recommend storage requirements;
 3. Preparation and expiration date;
 4. The date of receipt and date opened; and
 5. Other pertinent information required for proper use.

(Rule 1200-06-03-.09, continued)

- (d) Reagents, solutions, culture media, control materials, calibration materials and other supplies must be prepared, stored, and handled in a manner to ensure the following:
 - 1. Reagents, solutions, culture media, controls, calibration materials and other supplies are not used when they have exceeded their expiration date or when they have deteriorated or are of substandard quality. The laboratory must comply with the FDA product dating requirements of 21 CFR 610.53 for blood products and other biologicals, and with the labeling requirement of 21 CFR 809.10 for all other in-vitro diagnostics. Any exception to the product dating requirements in 21 CFR 610.53, will be granted by the FDA in the form of an amendment of the product license, in accordance with 21 CFR 610.53(d). All exceptions must be documented by the laboratory; and
 - 2. Components of reagent kits of different lot numbers are not interchanged unless otherwise specified by the manufacturer.
- (4) Standard: Procedure Manual.
 - (a) A written procedure manual for the performance of all analytical methods used by the laboratory must be readily available and followed by laboratory personnel. Textbooks may be used as supplements to these written descriptions but may not be used in lieu of the laboratory's written procedures for testing or examining specimens. Procedure should be substantially in compliance with the CLSI, GP-2A, current version, or any subsequent version.
 - (b) The procedure manual must include, when applicable to the test procedure, all of the following:
 - 1. Requirements for specimen collection and processing, and criteria for specimen rejection;
 - 2. Procedures for microscopic examination, including the detection of inadequately prepared slides;
 - 3. Step-by-step performance of the procedure, including test calculations and interpretation of results;
 - 4. Preparation of slides, solutions, calibrators, controls, reagents, stains and other materials using in testing;
 - 5. Calibration and calibration verification procedures;
 - 6. The reportable range for patient test results as established or verified in Rule 1200-06-03-.09(5);
 - 7. Control procedures;
 - 8. Remedial action to be taken when calibration or control results fail to meet the laboratory's criteria for acceptability;
 - 9. Limitations in methodologies, including interfering substances;
 - 10. Reference range (normal values);
 - 11. Imminent life-threatening laboratory results or "critical values";

(Rule 1200-06-03-.09, continued)

12. Current pertinent literature references;
 13. Appropriate criteria for specimen storage and preservation to ensure specimen integrity until testing is completed;
 14. The laboratory's system for reporting patient results including when appropriate, the protocol for reporting critical values;
 15. Description of the course of action to be taken in the event that a test system becomes inoperable; and
 16. Criteria for the referral of specimens including procedures for specimen submission and handling as described in Rule 1200-06-03-.07(1)(a).
- (c) Manufacturers' package inserts or operator manuals may be used, when applicable, to meet the requirements of parts (b)1. through (b) 13. of this paragraph. Information not provided by the manufacturer must be provided by the laboratory.
 - (d) Procedures must be re-approved, signed and dated by the medical laboratory director or designee.
 - (e) Procedures must be re-approved, signed and dated if the directorship of the laboratory changes.
 - (f) Each change in a procedure must be approved, signed, and dated by the current medical laboratory director or his or her designee approved, in writing, by the medical laboratory director.
 - (g) The laboratory must maintain a copy of each procedure with the dates of initial use and discontinuance. These records must be retained for two (2) years after a procedure has been discontinued.
- (5) Standard: Establishment and Verification of Method Performance Specifications. Prior to reporting patient test results, the laboratory must verify or establish, for each method, the performance specifications for the following performance characteristics: accuracy; precision; analytical sensitivity and specificity, if applicable; the reportable range of patient test results; the reference range(s) (normal values); and any other applicable performance characteristic.
- (a) The provisions of this section are not retroactive.
 - (b) A laboratory that introduces a new procedure for patient testing must, prior to reporting patient test results:
 1. Verify or establish for each method the performance specifications for the following performance characteristics, as applicable:
 - (i) Accuracy;
 - (ii) Precision;
 - (iii) Analytical sensitivity;
 - (iv) Analytical specificity to include interfering substances;
 - (v) Reportable range of patient test results;

(Rule 1200-06-03-.09, continued)

- (vi) Reference range(s); and
 - (vii) Any other performance characteristic required for test performance.
 - 2. Establish calibration and control procedures for patient testing as required under Rules 1200-06-03-.09(7) and 1200-06-03-.09(8), based upon the performance specifications verified or established in accordance with Rule 1200-06-03-.09(5).
 - (c) The laboratory must have documentation of the verification or establishment of all applicable test performance specifications.
- (6) Standard: Equipment Maintenance and Function Checks - The laboratory must perform equipment maintenance and function checks to assure accurate and reliable test results and reports. This includes electronic, mechanical and operational checks of equipment, instruments, and test systems necessary for proper test performance and test result reporting.
- (a) Maintenance of equipment, instruments, and test systems - For equipment, instruments or test systems the laboratory must:
 - 1. Establish a maintenance protocol that ensures equipment, instrument, and test system performance necessary for accurate and reliable test results and test result reporting;
 - 2. Perform maintenance with at least the frequency specified in the above-mentioned maintenance protocol; and
 - 3. Document all maintenance performed.
 - (b) Function checks of equipment, instruments, and test systems - For equipment, instruments, or test systems the laboratory must:
 - 1. Define a function check protocol that ensures equipment, instrument, and test system performance necessary for accurate and reliable test results and test result reporting;
 - 2. Perform function checks including background or baseline checks as specified in the above-mentioned function check protocol. Function checks must be within the laboratory's established limits before patient testing is conducted; and
 - 3. Document all function checks performed.
- (7) Standard: Calibration and Calibration Verification Procedures - Calibration and calibration verification procedures are required to substantiate the continued accuracy of the test method throughout the laboratory's reportable range for patient test results. Calibration is the process of testing and adjusting an instrument kit, or test system to provide a known relationship between the measurement response and the value of the substance that is being measured by the test procedure. Calibration verification is the assaying of calibration materials in the same manner as patient samples to confirm that the calibration of the instrument, kit, or test system has remained stable throughout the laboratory's reportable range for patient test results. The reportable range for patient test results is the range of test result values over which the laboratory can establish or verify the accuracy of instrument, kit or test system measurement response. Calibration and calibration verification must be performed and documented as required in this rule unless otherwise specified in paragraphs 1200-06-03-.09(12) through 1200-06-03-.09(34). For each method the laboratory must:

(Rule 1200-06-03-.09, continued)

(a) Perform calibration procedures

1. At a minimum, in accordance with manufacturer's instructions, if provided, using calibration materials provided or specified, as appropriate, and with at least the frequency recommended by the manufacturer;
2. In accordance with criteria established by the laboratory.
 - (i) Including the number, type and concentration of calibration materials, acceptable limits for calibration, and the frequency of calibration if manufacturer's instructions are not provided; and
 - (ii) Using calibration materials appropriate for the methodology and, if possible, traceable to a reference method or reference material of known value; and
3. Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification; and

(b) Perform calibration verification procedures

1. In accordance with the manufacturer's calibration verification instructions when they meet or exceed the requirements specified by the laboratory as described in Rule 1200-06-03-.09(7)(b)2., or
2. In accordance with criteria established by the laboratory
 - (i) Including the number, type, and concentration of calibration materials, acceptable limits for calibration verification and frequency of calibration verification; and
 - (ii) Using calibration materials appropriate for
 - (I) The methodology and, if possible, traceable to a reference method or reference material of known value; and
 - (II) Verifying the laboratory's established reportable range of patient test results, which must include at least a minimal (or zero[0]) value, a mid-point value, and a maximum value at the upper limit of that range; and
 - (iii) At least once every six (6) months and whenever any of the following occur:
 - (I) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not effect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes.

Note: If reagents are obtained from a manufacturer and all of the reagents for a test are packaged together, the laboratory is not required to perform calibration verification for each package of reagents, provided the packages of reagents are received in the same shipment and contain the same lot number;

(Rule 1200-06-03-.09, continued)

- (II) There is major preventive maintenance or replacement of critical parts that may influence test performance;
 - (III) Controls reflect an unusual trend or shift or are outside of the laboratory's acceptable limits and other means of assessing and correcting unacceptable control values have failed to identify and correct the problem; or
 - (IV) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification than specified in Rule 1200-06-03-.09(7)(b)2.(i), (ii), or (iii) and
- (c) Document all calibration verification procedures performed.
- (8) Standard: Control Procedures – Control procedures are performed on a routine basis to monitor the stability of the method or test system; control materials provide a means to indirectly assess the accuracy and precision of patient test results. Control procedures must be performed as defined in this paragraph unless otherwise specified in paragraphs 1200-06-03-.09 (12) through 1200-06-03-.09 (34).
- (a) Control samples must be tested in the same manner as patient specimens.
 - 1. For qualitative tests, the laboratory must include a positive and negative control with each run of patient specimens.
 - 2. For quantitative tests, the laboratory must include at least two (2) samples of different concentrations of either calibration materials, control materials, or a combination thereof with the frequency determined in subparagraph 1200-06-03-.09(3)(a), but not less frequently than once each run of patient specimens.
 - 3. For electrophoretic determinations
 - (i) At least one (1) control sample must be used in each electrophoretic cell; and
 - (ii) The control sample must contain fractions representative of those routinely reported in patient specimens.
 - 4. Each day of use, the laboratory must evaluate the detection phase of direct antigen systems using an appropriate positive and negative control material (organism or antigen extract). When direct antigen systems include an extraction phase, the system must be checked each day of use using a positive organism.
 - 5. If calibration materials and control materials are not available, the laboratory must have an alternative mechanism to assure the validity of patient test results.
 - (b) Control samples must be tested in the same manner as patient specimens.
 - (c) When calibration or control materials are used, statistical parameters (e.g. mean and standard deviation) for each lot number of calibration material and each lot of control material must be determined through repetitive testing.

(Rule 1200-06-03-.09, continued)

1. The stated values of an assayed control material may be used as the target values provided the stated values correspond to the methodology and instrumentation employed by the laboratory and are verified by the laboratory.
 2. Statistical parameters for unassayed materials must be established over time by the laboratory through concurrent testing with calibration materials or control materials having previously determined statistical parameters.
- (d) Control results must meet the laboratory's criteria for acceptability prior to reporting patient test results.
- (e) Reagent and supply checks.
1. The laboratory must check each lot or shipment of reagents, discs, stains, antisera and identification systems (systems using two (2) or more substrates) when prepared or opened for positive and negative reactivity, as well as graded reactivity if applicable.
 2. Each day of use (unless otherwise specified in this rule), the laboratory must test staining materials for intended reactivity to ensure predictable staining characteristics.
 3. The laboratory must check fluorescent stains for positive and negative reactivity each time of use (unless otherwise specified in this rule).
 4. The laboratory must check each batch or shipment of media for sterility, if it is intended to be sterile and if sterility is required for testing. Media must also be checked for its ability to support growth, and as appropriate, selectivity/inhibition and/or biochemical response. The laboratory may use manufacturer's control checks of media provided the manufacturer's product insert specifies that the manufacturer's quality control checks meet the current standards of the Clinical and Laboratory Standards Institute (CLSI) for media quality control. The laboratory must document that the physical characteristics of the media are not compromised and report any deterioration in the media to the manufacturer. The laboratory must follow the manufacturer's specifications for using the media and be responsible for the test results.
 5. A batch of media (solid, semi-solid, or liquid) consists of all tubes, plates, or containers of the same medium prepared at the same time and in the same laboratory, or, if received from an outside source of commercial supplier, consists of all of the plates, tubes or containers of the same medium that have the same lot numbers and are received in a single shipment.
- (f) For each method that is developed in-house, that is a modification of the manufacturer's test procedure, the laboratory must evaluate instrument and reagent stability and operator variance in determining the number, type, and frequency of testing calibration or control materials and establish criteria for acceptability used to monitor test performance during a run of patient specimen(s). A run is an interval within which the accuracy and precision of a testing system is expected to be stable, but cannot be greater than twenty-four (24) hours or less than the frequency recommended by the manufacturer. For each procedure, the laboratory must monitor test performance using calibration materials or a combination thereof.
- (9) Standard: Remedial Actions - Remedial action policies and procedures must be established by the laboratory and applied as necessary to maintain the laboratory's operation for testing patient specimens in a manner that assures accurate and reliable patient test results and

(Rule 1200-06-03-.09, continued)

reports. The laboratory must document all remedial actions taken when any of the following occur:

- (a) The laboratory shall take any remedial action to correct a noted error in specimen collection or loss of specimen.
 - (b) Test systems that do not meet the laboratory's established performance specifications, as determined in Rule 1200-06-03-.09(5) include but are not limited to the following:
 1. Equipment or methodologies that perform outside of established operating parameters or performance specifications;
 2. Patient test values that are outside of the laboratory's reportable range of patient test results; and
 3. The determination that the laboratory's reference range for a test procedure is inappropriate for the laboratory's patient population.
 - (c) Results of control and calibration materials that fail to meet the laboratory's established criteria for acceptability. All patient test results obtained in the unacceptable test run or since the last acceptable test run must be evaluated to determine if patient test results have been adversely affected. The laboratory must take the remedial action necessary to ensure the reporting of accurate and reliable patient test results;
 - (d) If the laboratory cannot report patient test results within its established time frames the laboratory must determine, based on the urgency of the patient test(s) requested, the need to notify the appropriate individual of the delayed testing; and
 - (e) Errors in the reported patient test results are detected. The laboratory must
 1. Promptly notify the authorized person ordering or individual utilizing the test results of reporting errors;
 2. Issue corrected reports promptly to the authorized person ordering the test or the individual utilizing the test results; and
 3. Maintain exact duplicates of the original report as well as the corrected report for two (2) years.
 - (f) There must be a procedure of review in operation to verify highly unusual results such as delta values and critical values. This review must be performed and documented by the supervisor or other person designated in writing on a daily basis to identify possibly erroneous tests.
- (10) Standard: Quality Control Records - The laboratory must document and maintain records of all quality control activities specified in paragraphs 1200-06-03-.09(2) through 1200-06-03-.09(34) and retain records for at least two (2) years. Immunohematology quality control records must be maintained for a period of no less than five (5) years. In addition, quality control records for blood and blood products must be maintained for a period not less than five (5) years. In addition, quality control records for blood and blood products must be maintained for a period not less than five (5) years after processing records have been completed, or six (6) months after the latest expiration date, whichever is the later date, in accordance with 21 CFR 6060.160(d).
- (11) Condition: Quality Control - Specialties and Subspecialties - The laboratory must establish and follow written policies and procedures for an acceptable quality control program that

(Rule 1200-06-03-.09, continued)

include verification and assessment of accuracy, measurement of precision and detection of error for all analyses and procedures performed by the laboratory. In addition to the general requirements specified in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11), the laboratory must meet the applicable requirements of paragraphs 1200-06-03-.09(12) through 1200-06-03-.09(34) or an CMS approved equivalent procedure and the following:

- (a) Meets quality control requirements specified in this paragraph; or
 - (b) Follows manufacturer's instructions when using products (instruments, kits, or test systems) as well as specialty and subspecialty quality control.
 - (c) Failure to meet any of the applicable conditions in paragraph 1200-06-03-.09(12) through 1200-06-03-.09(34) may result in revocation of licensure for the entire specialty or subspecialty to which the condition applies, in accordance with Rule 12-6-3-.05.
- (12) Condition: Microbiology - The laboratory must meet the applicable quality control requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and in paragraphs 1200-06-03-.09(13) through 1200-06-03-.09(17) for the subspecialties for which it is certified under the specialty of microbiology.
- (13) Condition: Bacteriology - To meet the quality control requirements for bacteriology in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and in paragraphs 1200-06-03-.09(13) through 1200-06-03-.09(17) for the subspecialties for which it is certified under the specialty of microbiology.
- (a) The laboratory must check positive and negative reactivity with control organisms as specified below:
 - 1. Each day of use for catalase, coagulase, beta-lactamase, and oxidase reagents and DNA probes;
 - 2. Each week of use for Gram and acid-fast stains, bacitracin, optochin, ONPG, X and V discs or strips; and
 - 3. Each month of use for antisera.
 - (b) Each week of use, the laboratory must check XV discs or strips with a positive control organism.
 - (c) For antimicrobial susceptibility tests, the laboratory must check each new batch of media and each lot of antimicrobial discs before, or concurrent with, initial use, using approved reference organisms.
 - 1. The laboratory's zone sizes or minimum inhibitory concentration for reference organisms must be within established limits before reporting patient results.
 - 2. The laboratory must use the appropriate control organism(s) to check the procedure as required by current law under CLIA.
- (14) Condition: Mycobacteriology - To meet the quality control requirements for mycobacteriology, the laboratory must comply with the applicable requirements in paragraph 1200-06-03-.09(1) through 1200-63-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
- (a) Each day of use, the laboratory must check the iron uptake test with at least one (1) acid-fast organism that produces a positive reaction and with an organism that

(Rule 1200-06-03-.09, continued)

- produces a negative reaction and check all other reagents or test procedures used for mycobacteria identification with at least one (1) acid-fast organism that produces a positive reaction.
- (b) The laboratory must check flurochrome acid-fast stains for positive and negative reactivity each week of use.
 - (c) The laboratory must check acid-fast stains each week of use with an acid-fast organism that produces a positive reaction.
 - (d) For susceptibility tests performed on *Mycobacterium tuberculosis*, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
- (15) Condition: Mycology - To meet the quality control requirements for mycology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) though 1200-06-03-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
- (a) Each day of use, the laboratory using the auxanographic medium for nitrate assimilation must check the nitrate reagent with a peptone control.
 - (b) Each week of use, the laboratory must check all reagents used with biochemical tests and other test procedures for mycological identification with an organism that produces a positive reaction.
 - (c) Each week of patient testing, the laboratory must check acid-fast stains for positive and negative reactivity.
 - (d) For susceptibility tests, the laboratory must test each drug each day of use with at least one (1) control strain that is susceptible to the drug. The laboratory must establish control limits. Criteria for acceptable control results must be met prior to reporting patient results.
- (16) Condition: Parasitology - To meet the quality control requirements for parasitology, the laboratory must comply with the applicable requirements of paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (c) of this paragraph. All quality control activities must be documented.
- (a) The laboratory must have available a reference collection of slides or photographs, and, if available, gross specimens for identification of parasites and must use these references in the laboratory for appropriate comparison with diagnostic specimens.
 - (b) The laboratory must calibrate and use the calibrated ocular micrometer for determining the size of ova and parasites, if size is a critical parameter. Recalibration must be performed if the ocular micrometer is moved to another microscope or objectives are changed.
 - (c) Each month of use, the laboratory much check permanent stains using a fecal sample control that will demonstrate staining characteristics.
- (17) Condition: Virology - To meet the quality control requirements for virology, the laboratory must comply with the applicable requirements in paragraph 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (c) of this paragraph. All quality control activities must be documented.

(Rule 1200-06-03-.09, continued)

- (a) The laboratory must have available host systems for the isolation of viruses and test methods for the identification of viruses that cover the entire range of viruses that are etiologically related to clinical diseases for which services are offered.
 - (b) The laboratory must maintain records that reflect the systems used and the reactions observed.
 - (c) In tests for the identification of viruses, the laboratory must simultaneously culture uninoculated cells or cell substrate control as a negative control to detect erroneous identification results.
- (18) Condition: Diagnostic Immunology - The laboratory must meet the applicable quality control requirements in paragraph 1200-06-03-.09(1) through 1200-06-03-.09(11) and paragraphs 1200-3-.09(19) through 1200-06-03-.09(20) for the subspecialties for which it is certified under the specialty of diagnostic immunology.
- (19) Condition: Syphilis Serology - To meet the quality control requirements for syphilis serology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (c) of this paragraph. All quality control activities must be documented.
- (a) For laboratories performing syphilis testing, the equipment, glassware, reagents, controls, and techniques for tests for syphilis must conform to manufacturers' specifications.
 - (b) The laboratory must run serologic tests on patient specimens concurrently with a positive serum control of known titer or controls of graded reactivity plus a negative control.
 - (c) The laboratory must employ positive and negative controls that evaluate all phases of the test system to ensure reactivity and uniform dosages.
 - (d) The laboratory may not report test results unless the predetermined reactivity pattern of the controls is observed.
 - (e) All facilities manufacturing blood and blood products for transfusion or serving as referral laboratories for these facilities must meet the syphilis serology testing requirements of 21 CFR 640.5(a).
- (20) Condition: General Immunology - To meet the quality control requirements for general immunology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
- (a) The laboratory must run serologic tests on patient specimens concurrently with a positive serum control of known titer or controls of graded reactivity and uniform dosages when positive and negative controls are not sufficient.
 - (b) The Laboratory must employ controls that evaluate all phases of the test system (antigens, complement, erythrocyte indicator systems, etc.) to ensure reactivity and uniform dosages when positive and negative controls alone are not sufficient.
 - (c) The laboratory may not report test results unless the predetermined reactivity pattern of the controls is observed.

(Rule 1200-06-03-.09, continued)

- (d) All facilities manufacturing blood and blood products for transfusion or serving as referral laboratories for these facilities must meet
 - 1. The HIV testing requirements of 21 CFR 610.45; and
 - 2. Hepatitis testing and requirements of 21 CFR 610.40.

- (21) Condition: Chemistry - The laboratory must meet the applicable quality control requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and paragraphs 1200-06-03-.09(22) through 1200-06-03-.09(24) for the subspecialties for which it is certified under the specialty of chemistry.

- (22) Condition: Routine Chemistry - To meet the quality control requirements for routine chemistry, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11). All quality control activities must be documented. In addition, for blood gas analyses, the laboratory must:
 - (a) Calibrate or verify calibration according to the manufacturer's instructions and with at least the frequency recommended by the manufacturer;
 - (b) Test one (1) sample of control material for each eight (8) hours of testing or as set out in the manufacturer's instructions;
 - (c) Use a combination of calibrators and control materials that include both low and high values on each day of testing or as set out in the manufacturer's instruction; and
 - (d) Include one (1) sample of calibration material or control material each time patients are tested unless automated instrumentation internally verifies calibration at least every thirty (30) minutes or as set out in the manufacturer's instruction.

- (23) Condition: Endocrinology - To meet the quality control requirements for endocrinology, the laboratory must comply with the applicable requirements contained in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11). All quality control activities must be documented.

- (24) Condition: Toxicology - To meet the quality control requirements for toxicology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11). However, this subsection shall not apply to drug testing for employment purposes. All quality control activities must be documented. In addition, for drug abuse screening using thin layer chromatography:
 - (a) Each plate must be spotted with at least one (1) sample of calibration material containing all drug groups identified by thin layer chromatography which the laboratory reports; and
 - (b) At least one (1) control sample must be included in each chamber, and the control sample must be processed through each step of patient testing, including extraction procedures.

- (25) Condition: Hematology - To meet the quality control requirements for hematology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
 - (a) For automated hematology testing systems, excluding coagulation, the laboratory must include two (2) levels of controls for each twenty-four (24) hours of operation.

(Rule 1200-06-03-.09, continued)

- (b) For all automated coagulation testing systems, the laboratory must include two (2) levels of control for each eight (8) hours of operation and each time a change in reagents occurs.
 - (c) For manual hemoglobin determinations (cyanmethemoglobin), the procedure must be standardized with reference materials of known, certified values. At least four (4) different hemoglobin concentrations must be used to prepare the calibration curve or to calibrate the readout instruments. For procedures using calibration curves, all curves must be repeated regularly and verified after serving or recalibration of instruments. Photometer functions checks must be run and recorded daily. Photometers must be checked for linearity periodically and when instrument adjustments have been made with appropriate filters or solutions.
 - (d) For manual hematocrit determinations, the speed and timer of the microhematocrit centrifuge must be checked at specific intervals. The constant packing time (minimum spin to reach maximum packing of cells) must be assessed on installation and reassessed when there has been a change in either the speed or time.
- (26) Condition: Pathology - The laboratory must meet the applicable quality control requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and paragraphs 1200-06-03-.09(27) through 1200-06-03-.09(29) for the subspecialties; for which it is certified under the specialty of pathology. All quality control activities must be documented.
- (27) Condition: Cytology - To meet the quality control requirements for cytology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and subparagraphs (a) through (g) of this paragraph.
- (a) The laboratory must assure that
 - 1. All gynecologic smears are stained using a Papanicolaou or modified Papanicolaou staining method;
 - 2. Effective measures are taken to prevent cross-contamination between gynecologic and nongynecologic specimens during the staining process;
 - 3. Nongynecologic specimens that have a high potential for cross-examination are stained separately from other nongynecologic specimens, and the stains are filtered or changed following staining;
 - 4. Diagnostic interpretations are not reported on unsatisfactory smears; and
 - 5. All cytology slide preparations are evaluated on the premises of a laboratory certified to conduct testing in the subspecialty of cytology.
 - (b) The laboratory is responsible for ensuring the following:
 - 1. Each individual engaged in the evaluation of cytology preparations by nonautomated microscopic technique examines no more than one hundred (100) slides (one patient per slide, gynecologic or nongynecologic, or both) in a twenty-four (24) hour period, irrespective of the site or laboratory. This limit represents an absolute maximum number of slides and is not being employed as a performance target for each individual. Previously examined reactive, reparative, atypical, premalignant or malignant gynecologic cases as defined in part (c)1. of this paragraph, previously examined nongynecologic cytology preparations, and tissue pathology slides examined by a pathologist are not included in the one hundred (100) slide limit.

(Rule 1200-06-03-.09, continued)

2. For purposes of workload calculations, each slide preparation (gynecologic or nongynecologic) made using automated, semi-automated, or other liquid-based slide preparatory techniques which result in cell dispersion over one-half ($\frac{1}{2}$) or less of the total available slide area and which is examined by nonautomated microscopic technique counts as one-half ($\frac{1}{2}$) slide.
3. Records are maintained of the total number of slides examined by each individual during each twenty-four (24) hour period, irrespective of the site or laboratory, and the number of hours each individual spends examining slides in the twenty-four (24) hour period; and
 - (i) The maximum number of one hundred (100) slides described in part (b)1. of this paragraph may not be examined in less than an eight (8) hour workday, and
 - (ii) For the purposes of establishing workload limits for individual examining slides by nonautomated microscopic technique on other than an eight (8) hour workday basis (includes full-time employees with duties other than slide examination and part-time employees), a period of eight (8) hours must be used to prorate the number of slides that may be examined. The following formula shall be used to determine maximum slide volume to be examined:
 - (iii) $(\text{Number of hours examining slides} \times 100) - 8$
- (c) A pathologist must ensure the following:
 1. All gynecologic smears interpreted to be showing reactive or reparative changes, atypical squamous or glandular cells of undetermined significance, or to be in the premalignant (dysplasia, cervical intraepithelial neoplasia or all squamous intraepithelial lesions including human papillomavirus-associated changes) or malignant category are confirmed by a pathologist qualified by certification in anatomic pathology by the American Board of Pathology. The report must be signed to reflect the review or, if a computer report is generated with signature, it must reflect an electronic signature authorized by the pathologist;
 2. All nongynecologic cytologic preparation are reviewed by the pathologist. The report must be signed to reflect the pathologist's review or, if a computer report is generated with signature, it must reflect an electronic signature authorized by the pathologist;
 3. The slide examination performance of each cytotechnologist is evaluated and documented, including performance evaluation through the reexamination of normal and negative cases and feedback on the reactive, reparative, a typical, malignant or premalignant cases as defined in part (c)1. of this paragraph; and
 4. A maximum number of slides, not to exceed the maximum workload limit described in subparagraph (b) of this paragraph is established by the pathologist for each individual examining slide preparations by nonautomated microscopic technique.
 - (i) The actual workload limit must be documented for each individual and established in accordance with the individual's capability based on the performance evaluation as described in part (c) 3. of this paragraph.

(Rule 1200-06-03-.09, continued)

- (ii) Records are available to document that each individual's workload limit is reassessed at least every six (6) months and adjusted when necessary.
- (d) The laboratory must establish and follow a program designed to detect errors in the performance of cytologic examinations and the reporting of results.
 - 1. The laboratory must establish a program that includes a review of slides from at least ten per cent (10%) of the gynecologic cases interpreted to be negative for reactive, reparative, atypical premalignant or malignant conditions as defined in part (c) 1. This review must be done by a pathologist qualified by certification in anatomic pathology by the American Board of Pathology, a cytology general supervisor qualified under Rule 1200-06-01-.23 or a cytotechnologist qualified under Rule 1200-06-01-.24.
 - (i) The review must include negative cases selected at random from the total caseload and from patients or groups of patients that are identified as having a high probability of developing cervical cancer, based on available patient information;
 - (ii) Records of initial examinations and rescreening results must be available; and
 - (iii) The review must be completed before reporting patient results on those cases selected.
 - 2. The laboratory must compare clinical information, when available, with cytology reports and must compare all malignant and premalignant (as defined in part (c)1. of this paragraph) gynecology reports with the histopathology report, if available in the laboratory (either on-site or in storage), and determine the causes of any discrepancies.
 - 3. For each patient with a current high grade intraepithelial lesion (moderate dysplasia or CIN-2 or above), the laboratory must review all normal or negative gynecologic specimens received within the previous five (5) years, if available in the laboratory (either on-site or in storage). If significant discrepancies are found that would affect patient care, the laboratory must notify the patient's physician and issue an amended report.
 - 4. The laboratory must establish and document an annual statistical evaluation of the number of cytology cases examined, number of specimens processed by specimen type, volume of patient cases reported by diagnosis (including the number reported as unsatisfactory for diagnostic interpretation), number of gynecologic cases where cytology and available histology are discrepant, the number of gynecologic cases where any rescreen of a normal or negative specimen results in reclassification as malignant or premalignant, as defined in part (c)1. of this paragraph, and the number of gynecologic cases for which history results were unavailable to compare with malignant or premalignant cytology cases as defined in part(c)1. of this paragraph.
 - 5. The laboratory must evaluate the case reviews of each individual examining slides against the laboratory's overall statistical values, document any discrepancies, including reasons for the deviation, and document corrective action, if appropriate.
- (e) The laboratory report must:

(Rule 1200-06-03-.09, continued)

1. Clearly distinguish specimens or smears, or both, that are unsatisfactory for diagnostic interpretation; and
 2. Contain narrative descriptive nomenclature for all results.
- (f) Corrected reports issued by the laboratory must indicate the basis for correction.
- (g) The laboratory must retain all slide preparations for five (5) years from the date of examination, or slides may be loaned to proficiency testing programs, in lieu of maintaining them for this time period, provided the laboratory receives written acknowledgement of the receipt of slides by the proficiency testing program and maintains the acknowledgement to document the loan of such slides. Documentation for slides loaned or referred for purposes other than proficiency testing must also be maintained. All slides must be retrievable upon request.
- (28) Condition: Histopathology - To meet the quality control requirements for histopathology, a laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and subparagraphs (a) through (c) of this paragraph. All quality control activities must be documented.
- (a) A control slide of known reactivity must be included with each slide or group of slides for differential, or special stains. Reaction(s) of the control slide with each special stain must be documented each day of testing.
 - (b) The laboratory performing the histopathology testing must retain stained slides at least ten (10) years from the date of examination and retain specimen blocks at least five (5) years from the date of examination.
 - (c) The laboratory must retain remnants of tissue specimens in a manner that assures proper preservation of the tissue specimens until the portions submitted for microscopic examination have been examined and a diagnosis made by an individual qualified by certification in anatomic pathology by the American Board of Pathology.
 - (d) All tissue pathology reports must be signed by an individual qualified as specified in subparagraph (c) of this paragraph. If a computer report is generated with an electronic signature, it must be authorized by the individual qualified as specified in subparagraph (c) of this paragraph.
 - (e) The laboratory must utilize acceptable terminology of a recognized system of disease nomenclature in reporting results.
- (29) Condition: Oral Pathology - To meet the quality control requirements for oral pathology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and 1200-06-03-.09(28). All quality control activities must be documented.
- (30) Condition: Radiobioassay - To meet quality control requirements for radiobioassay, the laboratory must comply with the applicable requirements of paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11). All quality control activities must be documented.
- (31) Condition: Histocompatibility - In addition to meeting the applicable requirements for general quality control in paragraphs 1200-06-03-.09(1), for quality control for general immunology in paragraph 1200-06-03-.09(20) and for immunohematology in paragraph 1200-06-03-.09(33), the laboratory must comply with the applicable requirements in subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.

(Rule 1200-06-03-.09, continued)

- (a) For renal allotransplantation, the laboratory must meet the following requirements:
1. The laboratory must have available and follow criteria for the following:
 - (i) Selecting appropriate patient serum samples for crossmatching;
 - (ii) The technique used in crossmatching;
 - (iii) Preparation of donor lymphocytes for crossmatching; and
 - (iv) Reporting crossmatch results;
 2. The laboratory must
 - (i) Have available results of final crossmatches before an organ or tissue is transplanted, and
 - (ii) Make a reasonable attempt and document efforts to have available serum specimens for all potential transplant recipients at initial typing, for periodic screening, for pretransplantation crossmatch and following sensitizing events, such as transfusion and transplant loss;
 3. The laboratory's storage and maintenance of both recipient sera and reagents must
 - (i) Be at an acceptable temperature range for sera and components;
 - (ii) Use a temperature alarm system and have an emergency plan for alternate storage; and
 - (iii) Ensure that all specimens are properly identified and easily retrievable;
 4. The laboratory's reagent typing sera inventory (applicable only to locally constructed trays) must indicate source, bleeding date and identification number, and volume remaining.
 5. The laboratory must properly label and store all labile components utilized in histocompatibility testing.
 6. The laboratory must:
 - (i) HLA type all potential transplant recipients;
 - (ii) Type cells from organ donors referred to the laboratory; and
 - (iii) Have available and follow a policy that establishes when antigen redefinition and retyping are required;
 7. The laboratory must have available and follow criteria for
 - (i) The preparation of lymphocytes for HLA-A, B and DR typing;
 - (ii) Selecting typing reagents, whether locally or commercially prepared;
 - (iii) The assignment of HLA antigens; and

(Rule 1200-06-03-.09, continued)

- (iv) Assuring that reagents used for typing recipients and donors are adequate to define all major and International Workshop HLA-A, B and DR specificities for which reagents are readily available;
8. The laboratory must do the following:
- (i) Screen potential transplant recipient sera for performed HLA-A and B antibodies with a suitable lymphocyte panel on sera collected
 - (I) At the time of the recipient's initial HLA typing; and
 - (II) Thereafter, following sensitizing events and upon request; and
 - (ii) Use a suitable cell panel for screening patient sera (antibody screen), a screen that contains all the major HLA specificities and common splits
 - (I) If the laboratory does not use commercial panels, it must maintain a list of individuals for fresh panel bleeding; and
 - (II) If the laboratory uses frozen panels, it must have a suitable storage system,
9. The laboratory must check
- (i) Each typing tray using
 - (I) Positive control sera;
 - (II) Negative control sera; and
 - (III) Positive controls for specific cell types when applicable (i.e., T. cells, B cells, and monocytes); and
 - (ii) Each compatibility test (i.e. mixed lymphocyte cultures, homozygous typing cells or DNA analysis) and typing for disease-associated antigens using controls to monitor the test components and each phase of the test system to ensure an acceptable performance level;
10. Compatibility testing for cellularly-defined antigens must utilize techniques such as the mixed lymphocyte culture test, homozygous typing cells or DNA analysis;
11. If the laboratory reports the recipient's or donor's, or both, ABO blood group and D(Rho) typing, the testing must be performed in accordance with 1200-06-03-.09(33);
12. If the laboratory utilizes immunologic reagents (such as antibodies or complement) to remove contaminating cells during the isolation of lymphocytes or lymphocyte subsets, the efficiency of the methods must be verified with appropriate quality control procedures;
13. At least once each month, the laboratory must have each individual performing tests evaluate a previously tested specimen as an unknown to verify his or her ability to reproduce test results. Records of the results for each individual must be maintained; and

(Rule 1200-06-03-.09, continued)

14. The laboratory must participate in at least one (1) national or regional cell exchange program, if available, or develop an exchange system with another laboratory in order to validate inter-laboratory reproducibility.
 - (b) If the laboratory performs histocompatibility testing for:
 1. Transfusion and other non-renal transplantation, excluding bone marrow and living transplant, all the requirements specified in this paragraph, as applicable, except for the performance of mixed lymphocyte cultures, must be met;
 2. Bone marrow transplantation, all the requirements specified in this paragraph, including the performance of mixed lymphocyte cultures or other augmented testing to evaluate class 11 compatibility, must be met; and
 3. Non-renal solid organ transplantation, the results of final crossmatches must be available before transplantation when the recipient has demonstrated presensitization by prior serum screening except for emergency situations. The laboratory must document the circumstances, if known, under which emergency transplants are performed, and records must reflect any information concerning the transplant provided to the laboratory by the patient's physician.
 - (c) Laboratories performing HLA typing for disease-associated studies must meet all the requirements specified in this paragraph except for the performance of mixed lymphocyte cultures, antibody screening and crossmatching.
 - (d) For laboratories performing organ donor HIV testing the requirements of paragraph 1200-06-03-.09(20) for the transfusion of blood and blood products must be met.
- (32) Condition: Clinical Cytogenetics - To meet the quality control requirements for clinical cytogenetics, the laboratory must comply with the applicable requirement of paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(11) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
 - (a) When determination of sex is performed by X and Y chromatin counts, these counts must be based on an examination of an adequate number of cells. Confirmatory testing such as full chromosome analysis must be performed for all atypical results.
 - (b) The laboratory must have records that reflect the media used and document the reactions observed, number of cells counted, the number of cells karyotyped, the number of chromosomes counted for each metaphase spread, and the quality of the banding that the resolution is sufficient to support the reported results; and that an adequate number of karyotypes are prepared for each patient.
 - (c) The laboratory also must have policies and procedures for assuring an accurate and reliable patient sample identification during the process of accessioning, cell preparation, photographing or other image reproduction technique, and photographic printing, and storage and reporting of results or photographs.
 - (d) The laboratory report must include the summary and interpretation of the observations and number of cells counted and analyzed and the use of recognized nomenclature, such as the International System of Cytogenetic Nomenclature.
 - (e) Adequate slides, films, hard copies of karyotypes and reports shall be retained for twenty (20) years.

(Rule 1200-06-03-.09, continued)

- (33) Condition: Immunohematology - To meet the quality control requirements for immunohematology, the laboratory must comply with the applicable requirements in paragraphs 1200-06-03-.09(1) through 1200-06-03-.09(II) and with subparagraphs (a) through (d) of this paragraph. All quality control activities must be documented.
- (a) The laboratory must perform ABO group and D(Rho) typing, unexpected antibody detection, antibody identification and compatibility testing in accordance with manufacturer's instruction, if provided, and as applicable, with 21 CFR 606 (with the exception of 21 CFR 606.20a, Personnel) and 21 CFR 640, et. seq.
 - (b) The laboratory must perform ABO group by concurrently testing unknown red cells with anti-A and anti-B grouping reagents. For confirmation of ABO group, the unknown serum must be tested with known A, and B red cells.
 - (c) The laboratory must determine the D(Rho) type by testing unknown red cells with anti-D(anti-Rho) blood grouping reagent.
 - (d) If required in the manufacturer's package insert for anti-D reagents, the laboratory must employ a control system capable of detecting false positive D(Rho) test results.
- (34) Condition: Transfusion services and bloodbanking - If a facility provides services for the transfusion of blood and blood products, the facility must be under the adequate control and technical supervision of a pathologist or other doctor of medicine or osteopathy meeting the qualifications in Rule 1200-6-1-.20, technical supervision in immunohematology. The facility must ensure that there are facilities for procurement, safekeeping and transfusion of blood and blood products and that blood and blood products must be available to meet the needs of the physicians responsible for the diagnosis, management, and treatment of patients. The facility meets this condition by complying with the standards in paragraphs 1200-06-03-.09(34)(a) through (f).
- (a) Standard: Immunohematological collection, processing, dating periods, labeling and distribution of blood and blood products.

In addition to the requirements of parts 1. through 4. of this subparagraph, the facility must also meet the applicable quality control requirements in 1200-06-03-.09(1) through 1200-06-03-.09(10).

- 1. Blood and blood product collection, processing and distribution must comply with 21 CFR Part 640 and 21CFR Part 606, and the testing laboratory must meet the applicable requirements of 42 CFR Part 493.
 - 2. Dating of blood and blood products must conform to 21 CFR 610.53.
 - 3. Labeling of blood and blood products conform to 21 CFR Part 606, Subpart G.
 - 4. Policies to ensure positive identification of a blood or blood product recipient must be established, documented and followed.
- (b) Standard: Blood and blood product storage facilities
- 1. The blood and blood products must be stored under appropriate conditions, which include an adequate temperature alarm system that is checked and documented daily.
 - (i) An audible alarm system, including a chart recorder, must monitor proper blood and blood product storage temperature continuously.

(Rule 1200-06-03-.09, continued)

- (ii) Activation, to include high and low limits, of alarm system must be documented periodically, no less frequently than quarterly.
 - (iii) A remote alarm system is required if location where blood or blood products are stored is not continuously staffed.
 - (iv) There must be a written procedure for response to an alarm when temperature limits are exceeded.
 - (v) There must be a written procedure for response to an alarm indicating specific personnel to respond and their appropriate action.
2. If blood is stored or maintained for transfusion purposes outside of a monitored refrigerator, the facility must ensure and document that storage conditions, including temperature, are appropriate to prevent deterioration of the blood or blood product.
- (c) Standard: Arrangement for services - In the case of services provided outside the blood bank, the facility must have an agreement reviewed and approved by the director that governs the procurement, transfer, and availability of blood and blood products.
 - (d) Standard: Provision of testing - There must be provision for prompt ABO blood group, D(Rho) type, unexpected antibody detection and compatibility testing in accordance with 1200-06-03-.09(34) of this rule and the laboratory investigation of transfusion reactions, either through the facility or under arrangement with an approved facility on a continuous basis, under the supervision of a pathologist or other doctor of medicine or osteopathy meeting the qualifications of 1200-06-01-.20.
 - (e) Standard: Retention of samples of transfused blood and patient specimens.
 - 1. Samples of each unit of transfused blood and the patient specimens must be retained for a minimum of seven (7) days for further testing in the event of reactions according to the facility's established procedures.
 - 2. The facility must promptly dispose of blood not retained for further testing that has passed its expiration date.
 - (f) Standard: Investigation of transfusion reactions.
 - 1. The facility, according to its established procedures, must promptly investigate all transfusion reactions occurring in all facilities for which it has investigational responsibility and make recommendations to the medical staff regarding improvements in transfusion procedures.
 - 2. The facility must document that all necessary remedial actions are taken to prevent future recurrence of transfusion reactions and that all policies and procedures are reviewed to assure that they are adequate to ensure the safety of individuals being transfused within the facility.
 - 3. When a fatality occurs as a result of a complication of blood or blood component transfusions, the Director, Office of Compliance and Biologics Quality, Center for Biological Evaluation and Research (CBER), FDA, must be notified within twenty-four (24) hours. Within seven (7) days after the fatality, a written report of the investigation must be submitted to the Director at 1401 Rockville Pike, Suite

(Rule 1200-06-03-.09, continued)

200 N, Rockville, MD 20852-1448. A copy of the report should be sent to the collecting facility if appropriate. See AABB Accreditations Requirements Manual.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-105, and 68-29-125. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed October 11, 2004; effective December 25, 2004. Amendment filed January 14, 2005; effective March 30, 2005. Amendment filed February 15, 2006; effective May 1, 2006. Amendment filed March 14, 2006; effective May 28, 2006.

1200-06-03-.10 PERFORMANCE IMPROVEMENT PROGRAM.

- (1) Condition: Performance improvement. Each laboratory must establish and follow written policies and procedures for a comprehensive quality assurance program which is designed to monitor and evaluate the ongoing and overall quality of the total testing process. The laboratory's performance improvement program must evaluate the effectiveness of its policies and procedures; identify and correct problems; assure the accurate, reliable and prompt reporting of test results; and assure the adequacy and competency of the staff.
 - (a) The laboratory must revise policies and procedures as necessary based upon the results of those evaluations.
 - (b) The laboratory must meet the standards of this rule as they apply to the services offered, tests performed, test results reported, and the unique practices of each testing entity. All performance improvement activities must be documented.
- (2) Standard: Patient test management assessment. The laboratory must have an ongoing mechanism for monitoring and evaluating the systems required under Rule 1200-06-03-.08, Patient Test Management. The laboratory must monitor, evaluate, and revise, if necessary, based on the results of its evaluations, the following:
 - (a) The criteria established for patient preparation, specimen collection, labeling, preservation and transportation;
 - (b) The information solicited and obtained on the laboratory's test requisition for its completeness, relevance, and necessity for testing of patient specimens;
 - (c) The use and appropriateness of the criteria established for specimen rejection;
 - (d) The completeness, usefulness, and accuracy of the test report information necessary for the interpretation or utilization of test results;
 - (e) The timely reporting of test results based on testing priorities (STAT, routine, etc.); and
 - (f) The accuracy and reliability of test reporting systems, appropriate storage of records and retrieval of test results.
- (3) Standard: Quality control assessment. The laboratory must have an ongoing mechanism to evaluate the corrective actions taken under paragraph 1200-06-03-.09(9), Remedial Actions. Ineffective policies and procedures must be revised based on the outcome of the evaluation. The mechanism must evaluate the effectiveness of corrective actions taken for the following:
 - (a) Problem identified during the evaluation of calibration and control data for each test, method;

(Rule 1200-06-03-.10, continued)

- (b) Problems identified during the evaluation of patient test values for the purpose of verifying the reference range of test method; and
 - (c) Errors detected in reported results.
- (4) Standard: Quality control assessment. Under Rule 1200-06-03-.07, Participation in Proficiency Testing, the corrective actions taken for any unsatisfactory or unsuccessful proficient testing result(s) must be evaluated for effectiveness.
- (5) Standard: Comparison of test results. If a laboratory performs the same test using different methodologists or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using different methodologies, instruments, or testing sites. If a laboratory performs tests that are not included under 42 CFR §493.901 of CLIA, Proficiency Testing Programs, the laboratory must have a system for verifying the accuracy and reliability of its test results are least twice a year.
- (6) Standard: Relationship of patient information to patients test results. The laboratory must have a mechanism to identify and evaluate patient test results that appear inconsistent with relevant criteria such as:
 - (a) Patient age;
 - (b) Sex;
 - (c) Diagnosis or pertinent clinical data, when provided;
 - (d) Distribution of patient test results when available; and
 - (e) Relationship with other test parameters, when available within the laboratory.
- (7) Standard: Personnel assessment. The laboratory must have an ongoing mechanism to evaluate the competency of all laboratory personnel
- (8) Standard: Communications.
 - (a) The laboratory must have a system in place to document problems that occur as a result of breakdowns in communication between the laboratory and the authorized individual who orders or receives the results of test procedures or examinations.
 - (b) Corrective actions taken to resolve the problems and to minimize communications breakdowns must be documented.
- (9) Standard: Complaint investigations. The laboratory must have a system in place to assure that all complaints and problems reported to the laboratory are documented. Investigations of complaints must be made, when appropriate, and, as necessary, corrective actions must be instituted.
- (10) Standard: Performance improvement review with staff. The laboratory must have a mechanism for documenting and assisting problems identified during performance improvement reviews and discussing them with the staff. The laboratory must take corrective actions that are necessary to prevent recurrence.
- (11) Standard: Performance improvement records. The laboratory must maintain documentation of all performance improvement activities including problems identified and corrective actions

(Rule 1200-06-03-.10, continued)

taken. All performance improvement records must be maintained for period of two (2) years and available to the Board or its designee upon request.

Authority: T.C.A. §§ 4-5-202, 4-5-204, and 68-29-105. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.11 FACILITIES, SAFETY, INFECTIOUS AND HAZARDOUS WASTE DISPOSAL.

- (1) Condition: The laboratory must have facilities, space and environmental conditions necessary for conducting the services offered, must provide a reasonably safe, working environment, and must properly dispose of infectious and hazardous wastes.
 - (a) Standard: Facilities. The laboratory must be constructed, arranged, and maintained to ensure the space, ventilation, utilities necessary for conducting all phases of testing, including the testing process, as appropriate.
 - (b) Standard: Safety. Safety precautions must be established, posted, and observed to ensure protection from physical hazards and biohazards materials. Adequate fire precautions shall be observed; volatile chemicals and flammable reagents must be properly stored.
 - (c) Infectious and Hazardous Waste Disposal. Each facility must develop, maintain, and implement written policies and procedures for the definition and handling of its infectious and hazardous wastes which must comply with the requirements of the Occupational Safety and Health Act, 29 U.S.C.A. §§ 65 1, et seq.
 - (d) The following waste shall be considered to be infectious waste:
 1. Waste contaminated by patients who are isolated due to communicable disease, as provided in the U.S. Centers for Disease Control "Guidelines for Isolation Precautions in Hospitals";
 2. Cultures and stocks of infectious agents including specimen cultures collected from medical and pathological laboratories, cultures and stocks of infectious agents from research and industrial laboratories, wastes from the production of biologicals, discarded live and attenuated vaccines, and culture dishes and devices used to transfer, inoculate, and mix cultures;
 3. Waste human blood and blood products such as serum, plasma, and other blood components;
 4. Pathological waste, such as tissues, organs, body parts, and body fluids that are removed during surgery and autopsy;
 5. All discarded sharps (e.g., hypodermic needles, syringes, pasture pipettes, broken glass, scalpel blades) used in patient care or which have come into contact with infectious agents during use in medical, research, or industrial laboratories;
 6. Contaminated carcasses, body parts, and bedding of animals that were exposed to pathogens in research, in the production of biologicals, or in the in vivo testing of pharmaceuticals;
 7. Other waste determined to be infectious by the facility in its written policy.

(Rule 1200-06-03-.11, continued)

- (2) Infectious and hazardous waste must be segregated from other waste at the point of generation (i.e., the point at which the material becomes a waste) within the facility.
- (3) Waste must be packaged in a manner that will protect waste handlers and the public from possible injury and disease that may result from exposure to the waste. Such packaging must provide for containment of the waste from the point of generation up to the point of proper treatment or disposal. Packaging must be selected and utilized for the type of waste the packaging will contain, how the waste will be treated and disposed, and how it will be handled and transported, prior to treatment and disposal.
 - (a) Contaminated sharps must be directly placed in leakproof, rigid, and puncture-resistant containers which must then be tightly sealed.
 - (b) Whether disposable or reusable, all containers, bags, and boxes used for containment and disposal of infectious waste must be conspicuously identified. Packages containing infectious waste which pose additional hazards (e.g., chemical, radiological) must also be conspicuously identified to clearly indicate those additional hazards.
 - (c) Reusable containers for infectious waste must be thoroughly sanitized each time they are emptied, unless the surfaces of the containers have been completely protected from contamination by disposable liners or other devices removed with the waste.
 - (d) Opaque packaging must be used for pathological waste.
- (4) After packaging, waste must be handled and transported by methods ensuring containment and preservation of the integrity of the packaging, including the use of secondary containment where necessary. Plastic bags of infectious waste must be transported by hand.
- (5) After packaging, waste must be handled and transported by methods ensuring containment and preserving the integrity of the packaging, including the use of secondary containment where necessary.
 - (a) Waste must not be compacted or ground (i.e., in a mechanical grinder) prior to treatment, except that pathological waste may be ground prior to disposal.
 - (b) Plastic bags of infectious waste must be transported by hand.

Authority: T.C.A. §§ 4-5-202, 4-5-204, and 68-29-105. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Amendment filed March 22, 2001; effective June 5, 2001. Repeal and new rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.12 REPORTABLE DISEASES.

The director of a medical laboratory shall submit reports and/or cultures of microorganisms of reportable diseases established by the Commissioner of Health to the Department in accordance with Tenn. Comp. R. & Regs. Chapter 1200-14-01.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-105, and 68-29-107. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed September 17, 2003; effective December 1, 2003. Amendment filed February 15, 2006; effective May 1, 2006. Amendment filed July 10, 2006; effective September 23, 2006. Amendment filed August 19, 2010; effective November 17, 2010; however on

(Rule 1200-06-03-.12, continued)

October 26, 2010 a 10-day stay was filed by the Government Operations Committee; new effective date October 27, 2010.

1200-06-03-.13 PERSONNEL POLICIES. Personnel policies, practices, and procedures that adequately support sound laboratory practice shall be available in written form. A current record shall be maintained on each employee and shall include evidence of current licensure, and a resume of training and experience, competency assessment, annual safety training and annual review of policies and procedures, as it pertains to the individual's job responsibilities. Employee health records may be maintained in a separate file.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Amendment filed April 25, 1985; effective May 25, 1985. Amendment filed July 13, 1990; effective August 27, 1990. Amendment filed February 26, 1991; effective April 12, 1991. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Amendment filed April 29, 2002; effective July 13, 2002. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Repeal and new rule filed October 3, 2014; effective January 1, 2015.

1200-06-03-.14 PERSONNEL REQUIREMENTS FOR A MEDICAL LABORATORY.

- (1) Condition: All laboratories regulated under T.C.A. 68, Chapter 29 shall meet the personnel standards in Chapter 1200-06-01.
- (2) Condition: Each laboratory shall designate a medical laboratory director, general supervisor and testing personnel who meet the qualifications of each level and fulfill the listed responsibilities for each level. If so qualified, one (1) individual may serve in more than one (1) position.
- (3) Condition: All individuals qualified to perform laboratory tests, including the medical laboratory director, shall possess a current license issued by the Department. Each individual shall be licensed at the highest level for which he qualified and functions. Laboratory procedures shall only be performed by appropriately licensed individuals, or personnel with a current trainee permit.
- (4) Standard: An adequate number of qualified testing personnel shall be provided for the volume and diversity of procedures and tests performed, as determined by the Department by on site survey.
- (5) A medical laboratory director may direct no more than three (3) laboratories unless authorized by the Board.
- (6) The medical laboratory director shall make periodic on-site visits at a minimum of once per month.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-105, 68-29-114, and 68-29-118. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Repeal filed May 3, 1995; effective July 17, 1995. New rule filed January 7, 1997; effective March 23, 1997. Repeal and new rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.15 SPECIAL REGULATIONS FOR ASTC, BLOOD DONOR CENTERS, AND PLASMAPHERESIS CENTERS.

- (1) An exception to the requirements of 1200-06-03-.14 concerning direction and supervision may be made for a licensed Ambulatory Surgical Treatment Center (ASTC) providing the following are met:

(Rule 1200-06-03-.15, continued)

- (a) Condition: The laboratory is directed by a physician currently licensed in Tennessee.
 - (b) Condition: The facility employs a licensed technologist as laboratory supervisor.
 - (c) Condition: The facility participates successfully in an approved proficiency testing program
 - (d) Condition: The facility is in substantial compliance with laboratory regulations as determined by an on-site survey conducted by a representative of the Department.
 - (e) Condition: The facility performs only the following procedures: CBC, urinalysis, ABO grouping and Rh typing, Pregnancy tests, Gram's stain, wet mounts and Rh immune globulin testing. Performance of additional procedures will necessitate that the laboratory meet all regulatory requirements.
- (2) An exception to the requirements of Rule 1200-06-03-.14 concerning direction, supervision and technical laboratory personnel may be made for Blood Donor Centers and Plasmapheresis Centers providing the following conditions are met:
- (a) The laboratory is directed by a physician licensed in Tennessee who possesses a current Tennessee license.
 - (b) Screening tests for the purpose of determining donor suitability are limited to:
 - 1. Hemoglobin or hematocrit.
 - 2. Total serum protein.
 - 3. Dipstick urine test for glucose and protein.
 - 4. Urine drug screens for opiates and morphine.
 - 5. Platelet pre-counts.
 - (c) Screening tests for educational purposes are limited to blood groups (ABO) and blood type (Rh).
 - (d) The personnel performing the screening tests shall be trained in the performance of these tests, quality control and recordkeeping duties. The Laboratory Director shall have on record at the facility where the "trained personnel" are employed, documentation of training with date(s) of training and signed by the person(s) administering the training. These "training personnel" shall not be required to have a medical laboratory personnel license.
 - (e) The facility is in compliance with Rules 1200-06-03-.01 through 1200-06-03-.11 and 1200-06-03-.13.
 - (f) The performance of additional testing that includes but is not limited to donor accessing (ABO, Rh, Antibody Detection and/or identification, STS, HIV, hepatitis tests, ALT, protein electrophoresis, etc.) shall require the facility to comply with Rule 1200-06-03-.14.
 - (g) The Donor Center/Plasmapheresis Center shall report the results of all positive tests for hepatitis to the Division of Communicable Disease Control of the Department within two (2) weeks after the last day of each month. The report shall contain the name,

(Rule 1200-06-03-.15, continued)

address, birth date and social security number of the donor and the name and address of the Center.

- (h) Any inquiries and/or complaints referred to the Department in regard to donor selection, bleeding products, etc. will be referred to the U.S. Food and Drug Administration for investigation.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed October 26, 1979; effective December 10, 1979. Amendment filed December 14, 1981; effective January 28, 1982. Amendment filed September 30, 1987; effective November 14, 1987. Repeal and new rule filed January 7, 1997; effective March 23, 1997. Amendment filed February 14, 2000; effective April 29, 2000. Amendment filed December 1, 2000; effective February 14, 2001. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed August 25, 2003; effective November 8, 2003. Amendments filed May 16, 2005; effective July 30, 2005.

1200-06-03-.16 ALTERNATE SITE TESTING.

(1) Point of Care Laboratory Testing.

- (a) Definition: "Point of Care" laboratory testing is laboratory testing performed by health care personnel/professionals not licensed by the Medical Laboratory Act, T.C.A. §§ 68-29-101, et seq., and is performed outside the duly licensed laboratory and under the auspices of a laboratory required to be licensed by the Department, pursuant to the Medical Laboratory Act.
- (b) All point of care laboratory testing must be approved by the Board in accordance with the following guidelines.
1. For those tests determined by the Board to be appropriate for safe and accurate performance in a point of care setting (see waived test list) written notification shall be submitted on forms provided by the Board to the Administrative Office of the Board.
 2. For all other tests not covered under part 1 of this subparagraph (above), a facility must petition, and appear before the Board and satisfy the Board that approval for point of care testing is appropriate. A facility that has received approval from the Board for instrument/method to be used by identified personnel and/or department(s) for point of care testing of an analyte does not have to appear before the Board again to add analytes to the approved instrument/method and may add them by submitting written notification to the Board's administrative office. It is only when the facility wants to change or add personnel or department(s) to the previously approve instrument/method that another hearing before the Board is required.
 3. Requests to perform non-waived point of care tests must be submitted in writing and must include at least:
 - (i) Statement of medical need for performance outside a laboratory setting.
 - (ii) Analyte and methodology.
 - (iii) Personnel authority and responsibility.
 - (iv) Quality assurance protocols.
 - (v) Maintenance of records.

(Rule 1200-06-03-.16, continued)

- (vi) Performance and improvement protocols.
- (c) The performance of point of care laboratory testing must comply with the following minimum guidelines:
1. Documentation must show that the method used for testing has been approved by the Medical Laboratory Director.
 2. Documentation must show that an adequate training protocol, demonstration of competency, and annual in-service with demonstration of competency for each person performing the testing, has been approved by the Medical Laboratory Director.
 3. Records shall be retained for at least two (2) years, which show that manufacturer recommendations regarding the performance of quality control were met. Quality control records shall be reviewed by a person with appropriate authority according to an established performance improvement program.
 4. Results of point of care laboratory testing must be recorded in an appropriate clinical record and include the identity of the patient, date and time of specimen collection, units of measurement, test location and identity of the analyst performing the test. Such records must be readily retrievable for inspection.
 5. The Medical Laboratory Director, once having established critical values, shall have the discretion to determine if, consistent with good patient care, there is a need for verification by the clinical laboratory when values fall above or below the established critical values.
 6. Written policies and procedures, as approved by the Medical Laboratory Director, shall be available in the laboratory and at all locations where point of care laboratory testing is performed. Such procedures shall be appropriate to the personnel and location performing the testing.
- (d) Approval to perform point of care laboratory testing may be granted annually provided all minimum guidelines are met. The criteria for withdrawal of approval to perform such testing shall include but not be limited to:
1. Performance of testing by unauthorized personnel.
 2. Failure to comply with the above stated guidelines
 3. Unsatisfactory performance on two (2) consecutive or two (2) of three (3) consecutive proficiency test surveys.
 4. Failure to comply with other State or Federal guidelines.
 5. Deficiencies without an acceptable plan of correction cited by the Tennessee Medical Laboratory Board surveyors upon and on-site inspection of the laboratory.
- (2) Physician's Office Laboratories - Physician Office Laboratories (POLs) are exempt from licensure requirements of the Medical Laboratory Act
- (a) To be eligible for this exemption, the following conditions must be met:

(Rule 1200-06-03-.16, continued)

1. The laboratory collects, accepts, and tests only specimens from the private and personal patients of the physician who operates the practice or from the private and personal patients of any physician who is a member of a medical/physician group practice that operates the laboratory regardless of the distance of any member physician's practice location from the group practice's laboratory or the number of specimens collected, accepted, and/or tested; and
 2. The laboratory must be operated by the physician or through the employees of the physician. In a medical/physician group practice, one (1) of the group's physicians must be designated to operate the laboratory. The designated physician is responsible for actual supervision and direct responsibility for the performance of the laboratory and its personnel which includes, but is not limited to, actual supervision and direct responsibility for quality assurance, quality control, and test management; and
 3. The tests performed in the laboratory are used only for diagnosis and/or treatment of patients of the individual or group practice and are maintained in the practice's medical records for the patients for whom the tests were performed.
- (b) In the case of a medical/physician group practice, proof of affiliation with the group practice must be maintained at all offices in which the laboratory is not physically located and produced upon request by an authorized agent of the Department.
- (c) Industrial or company physician practices, student health services and other arrangements in which a licensed physician is responsible for the continuing care of a group of patients on an ongoing basis will be designated to be POLs.
- (3) Screening Programs – Screening programs are offerings of specified medical laboratory tests to the general public, the purpose of which is educational rather than for diagnosis of disease, and the results of which are immediately available on the site of the program to the person being tested except for those tests which for methodological reasons must be submitted to a medical laboratory in which case they shall be sent to a Tennessee licensed medical laboratory.
- (a) Screening programs conducted by for-profit hospitals or nonprofit organizations are exempt from the licensure requirements of the Medical Laboratory Act, pursuant to T.C.A. § 68-29-104(6), when the following conditions are met:
- (b) The screening program must be under the direct supervision of a physician licensed in Tennessee. Direct supervision means that a physician will be responsible for quality assurance of the testing performed and review of the results of such testing. The physician is not required to be on site for the screening program event.
- (c) The results of the screening program testing must be submitted to the personal physician of the individual being screened or reviewed by the physician responsible for the screening program. During the screening process, the individual being screened must be afforded reasonable privacy and, when required, on-site confidential counseling about the results of the testing.
- (d) A written notification to conduct a screening program must be submitted to the Administrative Office for the Medical Laboratory Board, indicating compliance with these rules and providing the following information:
1. Location of testing.
 2. Date of testing.

(Rule 1200-06-03-.16, continued)

3. Type of tests to be performed and methodology to be used.
 4. Name of the licensed physician supervising the screening.
 5. Name and address of the for-profit hospital or nonprofit organization conducting the screening.
 6. Statement indicating that the testing staff has appropriate training and competency to conduct the testing.
 7. Assurance that the for-profit hospital or nonprofit organization is in compliance with Rule 1200-06-03-.11 regarding the handling of infectious and hazardous waste.
- (e) A copy of the written notification submitted to the Administrative office must be retained by the for-profit hospital or nonprofit organization conducting the screening program and must be available for inspection at the site of the screening program.
- (f) Notification of screening is not required when the for-profit hospital or nonprofit organization is a licensed medical laboratory, provided the laboratory has complied with Rules 1200-06-03-.17 regarding the performance of waived testing and 1200-06-03-.16 regarding point of care testing.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed February 14, 2000; effective April 29, 2000. Amendment filed December 1, 2000; effective February 14, 2001. Repeal and new rule filed June 18, 2002; effective September 1, 2002. Amendment filed May 23, 2003; effective August 6, 2003. Amendment filed February 15, 2006; effective May 1, 2006. Amendment filed March 14, 2006; effective May 28, 2006. Amendment filed July 10, 2006; effective September 23, 2006. Amendment filed December 27, 2006; effective March 12, 2007. Amendment filed October 3, 2014; effective January 1, 2015.

1200-06-03-.17 WAIVED TESTING.

- (1) Definition: "Waived" means those laboratory tests, as defined by the Board, which may be performed by individuals not licensed under the Medical Laboratory Act, and which pose no reasonable risk of harm if performed incorrectly.
- (2) Clinical laboratory tests which are designated as waived shall:
 - (a) be determined by the Board;
 - (b) be an analyte and/or methodology previously deemed as waived under CLIA; and
 - (c) apply to any and all testing sites
- (3) Clinical laboratories, pharmacies, and other health care facilities licensed by the State of Tennessee who perform waived testing must comply with the following minimum guidelines:
 - (a) Notification must be provided to the Administrative Office for the Board using Board approved forms and protocols.
 - (b) Performance of waived testing by a licensed medical laboratory must be approved by a Medical Laboratory Director as defined by the Tennessee Medical Laboratory Act, T.C.A. § 68-29-103.

(Rule 1200-06-03-.17, continued)

- (c) In the absence of a Medical Laboratory Director as defined by the Tennessee Medical Laboratory Act, T.C.A. § 68-29-103, performance of waived testing must be approved by the physician legally responsible for laboratory testing performed by that health care facility or pharmacy.
- (d) Personnel must perform such testing in accordance with manufacturer recommendations, including quality control.
- (e) Patient test results must be recorded in a clinical record, to include date and time of specimen collection, units of measurement and identity of analyst performing the test.
- (f) The laboratory, pharmacy, or health care facility must identify personnel responsible for performing and supervising the waived testing.
- (g) Personnel performing the waived testing must have adequate, specific training and orientation to perform the tests, and must demonstrate satisfactory levels of competence before performing patient testing, and at a minimum, annually thereafter.
- (h) Performance of waived testing will be evaluated by Laboratory surveyors at the time of inspection.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.18 COLLECTION STATIONS.

- (1) Facilities deemed as collection stations shall perform no clinical testing procedures on site.
- (2) Collection stations owned and operated by a clinical laboratory licensed under the Tennessee Medical Laboratory Act shall not be required to obtain an individual license.
- (3) Collections must submit to the requirements set forth in Rule 1200-06-03-.08, .09, .10, .13, and .14.
- (4) On-site inspections shall be performed per Rule 1200-06-03-.04.
- (5) It shall be the responsibility of the owner to notify the Department in writing of a change in the location or physician or owner within fifteen (15) days of the actual change. A replacement license shall be issued by the Department at no cost for the remainder of the calendar year for a new location, director, or owner, provided that in the case of change of directors, the new director is already licensed or is eligible for licensure as a medical laboratory director under the Medical Laboratory Act and the regulations promulgated thereunder.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-104, and 68-29-105. **Administrative History:** Original rule filed June 18, 2002; effective September 1, 2002.

1200-06-03-.19 PREPARATORY PORTIONS OF LABORATORY TESTS.

- (1) Tennessee Code Annotated, Section 68-29-129 makes it a violation of the law to allow any person to perform "medical laboratory procedures" except individuals licensed and registered under this chapter. For purposes of that section, the words "medical laboratory procedures" include only the non-preparatory portions of any test required to be performed in a medical laboratory. The preparatory portions of the tests, while they must be performed in a facility licensed by the Board, do not require the extensive education and specialized knowledge required for licensure and may therefore be performed by persons not licensed under the Medical Laboratory Act.

(Rule 1200-06-03-.19, continued)

- (2) The following are considered to be preparatory portions of tests to be performed in a medical laboratory that may be assigned to persons not licensed under the Medical Laboratory Act:
 - (a) centrifuging, pouring off and preparing specimens for testing.
 - (b) preparing peripheral smears at bedside or in the laboratory.
 - (c) staining peripheral smears by automated methods.
 - (d) loading primary bar-coded specimens on analyzers.
 - (e) automated process of sorting, decapping, aliquoting and archiving of specimens.
 - (f) primary inoculation of microbiology specimens.
 - (g) Any activities required prior to microscopic evaluation of cytology specimens.
- (3) A medical laboratory director, as defined by T.C.A. § 68-29-103, must approve preparatory portions of tests performed by individuals not licensed under the Tennessee Medical Laboratory Act.
- (4) The laboratory must identify personnel responsible for performing preparatory portions of tests and those with responsibility for supervising them.
- (5) Personnel performing preparatory portions of tests must have adequate, specific training and orientation and must demonstrate satisfactory levels of competency before performing preparatory portions of tests, and a competency demonstration must be performed at least annually thereafter.
- (6) Laboratory surveyors will evaluate preparatory portions of tests at the time of inspection.

Authority: §§ 4-5-202, 4-5-204, 68-29-105, and 68-29-129. **Administrative History:** Original rule filed April 10, 2003; effective June 24, 2003. Amendment filed September 17, 2003; effective December 1, 2003. Amendment filed April 17, 2007; effective July 1, 2007.

1200-06-03-.20 ADVERTISING.

- (1) Policy Statement. The lack of sophistication on the part of many in the health care community concerning medical laboratories, the importance of the interests affected by the choosing of a medical laboratory and the foreseeable consequences of unrestricted advertising by medical laboratories which is recognized to pose special possibilities for deception, require that special care be taken by medical laboratories to avoid misleading the health care community. Medical laboratories must be mindful that the benefits of advertising depend upon its reliability and accuracy. Since advertising by medical laboratories is calculated and not spontaneous, reasonable regulation designed to foster compliance with appropriate standards serves the public interest without impeding the flow of useful, meaningful, and relevant information to the health care community.
- (2) Definitions
 - (a) Advertisement – Informational communication to the health care community in any manner designed to attract attention to the medical laboratories which are licensed to practice in Tennessee.

(Rule 1200-06-03-.20, continued)

- (b) Licensee - Any entity holding a license to operate as a medical laboratory in the State of Tennessee. Where applicable this shall include partnerships and/or corporations.
 - (c) Material Fact - Any fact which a health care provider would need to know or rely upon in order to make an informed decision concerning the choice of medical laboratories to serve its particular needs.
 - (d) Bait and Switch Advertising - An alluring but insincere offer to sell a product or service which the advertiser in truth does not intend or want to sell. Its purpose is to switch consumers from buying the advertised service or merchandise, in order to sell something else, usually for a higher fee or on a basis more advantageous to the advertiser.
 - (e) Discounted Fee - Shall mean a fee offered or charged by a medical laboratory or a product or service that is less than the fee the medical laboratory usually offers or charges for the product or service. Products or services expressly offered free of charge shall not be deemed to be offered at a "discounted fee."
 - (f) Health Care Community – Shall mean hospitals, ambulatory surgical treatment centers, medical practices, individual physicians, and other health care providers with legal authority to order laboratory tests.
- (3) Advertising Fees and Services
- (a) Fixed Fees - Fixed fees may be advertised for any service. It is presumed unless otherwise stated in the advertisement that a fixed fee for a service shall include the cost of all professional recognized components within generally accepted standards that are required to complete the service.
 - (b) Range of Fees. A range of fees may be advertised for services and the advertisement must disclose the factors used in determining the actual fee, necessary to prevent deception of the health care community.
 - (c) Discount Fees. Discount fees may be advertised if:
 - 1. The discount fee is in fact lower than the licensee's customary or usual fee charged for the service; and
 - 2. The licensee provides the same quality and components of service and material at the discounted fee that are normally provided at the regular, non-discounted fee for that service.
 - (d) Related Services and Additional Fees. Related services which may be required in conjunction with the advertised services for which additional fees will be charged must be identified as such in any advertisement.
 - (e) Time Period of Advertised Fees.
 - 1. Advertised fees shall be honored for those seeking the advertised services during the entire time period stated in the advertisement whether or not the services are actually rendered or completed within that time.
 - 2. If no time period is stated in the advertisement of fees, the advertised fee shall be honored for thirty (30) days from the last date of publication or until the next scheduled publication whichever is later whether or not the services are actually rendered or completed within that time.

(Rule 1200-06-03-.20, continued)

- (4) Advertising Content. The following acts or omissions in the context of advertisement by any licensee shall constitute unethical conduct, and subject the licensee to disciplinary action pursuant to T.C.A. § 68-29-127(9) and (10).
- (a) Claims that the services performed, personnel employed, materials or office equipment used are professionally superior to that which is ordinarily performed, employed, or used, or that convey the message that one licensee is better than another when superiority of services, personnel, materials or equipment cannot be substantiated.
 - (b) The misleading use of an unearned or non-health degree in any advertisement.
 - (c) Promotion of professional services which the licensee knows or should know is beyond the licensee's ability to perform.
 - (d) Techniques of communication which intimidate, exert undue pressure or undue influence over a prospective client.
 - (e) Any appeals to an individual's anxiety in an excessive or unfair manner.
 - (f) The use of any personal testimonial attesting to a quality of competency of a service or treatment offered by a licensee that is not reasonably verifiable.
 - (g) Utilization of any statistical data or other information based on past performances for prediction of future services, which creates an unjustified expectation about results that the licensee can achieve.
 - (h) The communication of personal identifiable facts, data, or information about a patient without first obtaining patient consent.
 - (i) Any misrepresentation of a material fact.
 - (j) The knowing suppression, omission or concealment of any material fact or law without which the advertisement would be deceptive or misleading.
 - (k) Statements concerning the benefits or other attributes of medical procedures or products that involve significant risks without including:
 - 1. A realistic assessment of the safety and efficiency of those procedures or products; and
 - 2. The availability of alternatives; and
 - 3. Where necessary to avoid deception, descriptions or assessment of the benefits or other attributes of those alternatives.
 - (l) Any communication which creates an unjustified expectation concerning the potential results of any laboratory test.
 - (m) Failure to comply with the rules governing advertisement of fees and services, or advertising records.
 - (n) The use of "bait and switch" advertisements. Where the circumstances indicate "bait and switch" advertising, the Board may require the licensee to furnish data or other evidence pertaining to those sales at the advertised fee as well as other sales.

(Rule 1200-06-03-.20, continued)

- (o) Misrepresentation of a licensee's credentials, training, experience, or ability.
 - (p) Failure to include the corporation, partnership or individual licensee's name, address, and telephone number in any advertisement. Any corporation, partnership or association which advertises by use of a trade name or otherwise fails to list all licensed laboratory personnel practicing at a particular location shall:
 - 1. Upon request provide a list of all licensed laboratory personnel practicing at that location; and
 - 2. Maintain and conspicuously display at the licensee's office, a directory listing all licensed laboratory personnel practicing at that location.
 - (q) Failure to disclose the fact of giving compensation or anything of value to representatives of the press, radio, television or other communicative medium in anticipation of or in return for any advertisement (for example, newspaper article) unless the nature, format or medium of such advertisement make the fact of compensation apparent.
 - (r) After thirty (30) days of the licensee's departure, the use of the name of any licensed laboratory personnel formerly practicing at or associated with any advertised location or on office signs or buildings. This rule shall not apply in the case of a retired or deceased former associate who practiced in association with one or more of the present licensees if the status of the former associate is disclosed in any advertisement or sign.
 - (s) Stating or implying that a certain licensee provides all services when any such services are performed by another licensee.
- (5) Advertising Records and Responsibility
- (a) Each licensee who is a principal partner, or officer of a firm or entity identified in any advertisement, is jointly and severally responsible for the form and content of any advertisement. This provision shall also include any licensed professional employees acting as an agent of such firm or entity.
 - (b) Any and all advertisement are presumed to have been approved by the licensee named therein.
 - (c) A recording of every advertisement communicated by electronic media, and a copy of every advertisement communicated by print media, and a copy of any other form of advertisement shall be retained by the licensee for a period of two (2) years from the last date of broadcast or publication and be made available for review upon request by the Board or its designee.
 - (d) At the time any type of advertisement is placed, the licensee must possess and rely upon information which, when produced, would substantiate the truthfulness of any assertion, omission or representation of material fact set forth in the advertisement or public information.
- (6) Use of Titles in Advertisements. Any medical laboratory licensed by the Board which includes in its advertisements the names of its laboratory personnel must, in every "advertisement" [as that term is defined in rule 1200-06-03-.20(2)(a)] it publishes, use an appropriate personnel title for each such licensee as authorized by rule 1200-06-01-.03(3) of this rule. Failure to do so may constitute an omission of a material fact which makes the advertisement misleading and deceptive and subjects the medical laboratory to disciplinary action pursuant to T.C.A. §

(Rule 1200-06-03-.20, continued)

68-29-126(9), 68-29-126(11), 68-29-126(13), 68-29-127(8), 68-29-127(9), 68-29-127(10), and/or 68-29-129(8).

- (7) Severability. It is hereby declared that the sections, clauses, sentences and parts of these rules are severable, are not matters of mutual essential inducement, and any of them shall be rescinded if these rules would otherwise be unconstitutional or ineffective. If any one or more sections, clauses, sentences or parts shall for any reason be questioned in court, and shall be adjudged unconstitutional or invalid, such judgment shall not affect, impair or invalidate the remaining provisions thereof, but shall be confined in its operation to the specific provision or provisions so held unconstitutional or invalid, and the inapplicability or invalidity of any section, clause, sentence or part in any one or more instance shall not be taken to affect or prejudice in any way its applicability or validity in any other instance.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 63-1-145, 63-1-146, 68-29-105, 68-29-126, and 68-29-129.
Administrative History: Original rule filed March 16, 2007; effective May 30, 2007. Amendment filed April 22, 2010; effective July 21, 2010.

1200-06-03-.21 RENEWAL OF LABORATORY LICENSE.

- (1) The license to operate a laboratory or collection station shall expire annually on the anniversary of the date that the license was originally issued. To continue operations, laboratories and collection stations must renew their licenses on or before the expiration date.
- (2) Methods of Renewal
1. Internet Renewals - Laboratories and collection stations may apply for renewal and pay the necessary fees via the Internet. The application to renew can be accessed at:

www.tennessee.gov/health
 2. Paper Renewals - For laboratories or collection stations that have not renewed their license online via the Internet, a renewal application form will be mailed to each laboratory and collection station licensed by the Board at least sixty (60) days prior to the expiration date to the last address provided to the Board. Failure to receive such notification does not relieve the laboratory or collection station from the responsibility of meeting all requirements for renewal.
- (3) To be eligible for renewal, a laboratory or collection station must submit to the Division of Health Related Boards on or before the expiration date all of the following:
1. A completed renewal application form, signed by the owner and director of the medical laboratory or public official responsible for the operation of a city or county medical laboratory or institution that contains a medical laboratory; and
 2. In alphabetical order, a roster of all personnel currently employed in the laboratory, the classification/category in which the employee functions and is licensed, license number, expiration date and social security number. This shall include all medical laboratory directors, consultants, supervisors and testing personnel; and
 3. The renewal and state regulatory fees as provided in Rule 1200-06-03-.02.
- (4) All laboratories performing tests must file a separate renewal application for each laboratory address.

(Rule 1200-06-03-.21, continued)

- (5) Laboratories within a hospital that are located in contiguous buildings on the same campus and under common direction may file a single renewal application or multiple renewal applications for the laboratory sites within the same physical location or street address.
- (6) Laboratories and collection stations which submit a signed renewal form or letter that is found to be untrue may be subject to disciplinary action as provided in Rule 1200-06-03-.05.
- (7) Laboratories and collection stations which fail to comply with the renewal rules or notification received by them concerning failure to timely renew shall have their licenses processed pursuant to rule 1200-10-01-.10.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-105, 68-29-113, and 68-29-126. **Administrative History:** Original rule filed April 17, 2007; effective July 1, 2007.

1200-06-03-.22 SPECIAL REGULATIONS REGARDING LABORATORIES LICENSED UNDER T.C.A. § 68-29-138.

- (1) Medical laboratories performing advanced esoteric applied toxicological, forensic, or biochemical analysis utilizing emerging technology shall be licensed.
 - (a) Licensing procedures set out in Rule 1200-06-03-.02 shall be met.
 - (b) Laboratories licensed under T.C.A. § 68-29-138(a) are subject to all provisions of the Medical Laboratory Act except for the provisions of the Medical Laboratory Act and rules promulgated under the Act that require persons who accept specimens for laboratory examination and perform analytical testing or report the results of a laboratory examination to be licensed as a technologist, technician, laboratory trainee, or special analyst.
 1. To be eligible for the licensed personnel exemption, a medical laboratory must obtain a license pursuant to T.C.A. § 68-29-138. This license shall be unique and in addition to other licenses for laboratories located in contiguous buildings on the same campus that do not fall within the exemption.
 - (c) The application for licensure under this exemption shall require a specific testing menu.
- (2) Laboratories licensed under this rule shall be subject to on-site surveys by the Department. The survey shall include inspection of the following: laboratories, tools therein, and educational requirements of personnel defined in T.C.A. § 68-29-138(c).
- (3) Only medical laboratories engaged in advanced esoteric applied toxicological, forensic, or biochemical laboratory analysis utilizing emerging technologies shall be given an exemption defined in subparagraph (1)(b).
 - (a) Definition: advanced esoteric applied toxicological, forensic, or biochemical laboratory analysis is narrowly focused analysis that takes place within independent laboratories.
 - (b) Emerging technologies include, but are not limited to, chromatographic and non-chromatographic techniques coupled with mass spectrometer based detector systems and molecular diagnostic techniques.
- (4) Personnel in laboratories licensed under T.C.A. § 68-29-138 shall not be required to complete the procedures for personnel licensure as provided in Rule 1200-06-01-.05. However, personnel who analyze or report laboratory results shall possess a baccalaureate

(Rule 1200-06-03-.22, continued)

or advanced degree in the appropriate sciences with at least sixteen (16) semester hours of chemistry and/or biological science courses.

- (a) Licensed personnel shall be allowed to perform tasks for which they are appropriately licensed and trained in laboratories exempt from hiring licensed personnel.
- (b) Determining the appropriate sciences for baccalaureate or advanced degree for unlicensed personnel shall be left to the medical director of the laboratory.

Authority: T.C.A. §§ 4-5-202, 4-5-204, 68-29-103, 68-29-104, 68-29-105, 68-29-111, 68-29-113, 68-29-114, 68-29-115, and 68-29-138. *Administrative History:* Original rule filed April 1, 2015; effective June 30, 2015.